

HUMAN SERUM ALBUMIN ELISA KIT

FOR THE QUANTITATIVE DETERMINATION
OF HUMAN ALBUMIN CONCENTRATIONS
IN SERUM AND EDTA PLASMA



ALWAYS REFER TO LOT SPECIFIC PROTOCOL
PROVIDED WITH EACH KIT FOR
INSTRUCTIONS. PROTOCOL MUST BE
READ BEFORE USING THIS PRODUCT.

FOR RESEARCH USE ONLY. NOT FOR USE IN
DIAGNOSTIC PROCEDURES.

PRODUCT INFORMATION:

THIS KIT IS FOR ONE TIME USE ONLY.

ELISA NAME	HUMAN SERUM ALBUMIN ELISA
Catalog No.	SK00383-06
Lot No.	
Formulation	96 T
Standard Range	1.56 -100 ng/mL
Sensitivity	0.3 ng/mL
Sample Require	5~ 10 μ L
Dilution Factor	1,000,000~4,000,000 (1000K~4000K) (Optimal dilutions should be determined by each laboratory for each application)
Sample Type	Serum and EDTA Plasma
Specificity	Human Albumin
Calibration	Human Albumin
Intra-assay Precision	4 - 8%
Inter-assay Precision	8 - 12%
Storage	2 – 8°C
This kit contains sufficient materials to run 35 samples duplicated provided that assay is run according to protocol.	

Order Contact:
AVISCIERA BIOSCIENCE, INC.
2348 Walsh Ave., Suite C
Santa Clara, CA 95051
USA

Tel: (408) 982 0300

Fax: (408) 982 0301

Email: Sales@AvisceraBioscience.com

Info@AvisceraBioscience.com

www.AvisceraBioscience.com

DESCRIPTION

This Human Albumin ELISA Kit contains the necessary components required for the quantitative measurement of natural Human albumin from serum and EDTA plasma in a sandwich ELISA format.

This immunoassay contains Human albumin and antibodies raised against this protein. Results from this immunoassay have shown to accurately quantify natural Human albumin samples.

ASSAY OVERVIEW

The Human Albumin ELISA kit is based on the binding of Human albumin in samples to two antibodies. One monoclonal antibody has been pre-coated onto a microplate, and the other polyclonal antibody. Standards and samples are pipetted into the wells and any albumin present is bound by the immobilized antibody. After a washing step, the anti rabbit antibody-HRP conjugate is added to the wells. After washing away any unbound enzyme, a substrate solution is added to the wells and color develops in proportion to the amount of albumin bound in the initial step. The color development is stopped and the intensity of the color is measured.

PROCEDURAL LIMITATIONS

_FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.

_This ELISA kit should not be used beyond the expiration date on the kit label.

_Do not mix reagents with those from other lots or sources.

_It is important that the Dilution Buffer selected for the standard curve be consistent with the samples being assayed.

_Each laboratory must determine the optimal dilution factors for the samples being assayed with a pretest. If samples generate values that are not within the dynamic range of the standard curve, further concentrate or dilute the samples as required with Dilution Buffer and repeat the assay.

_Any modifications in buffers, pipetting technique, washing technique, incubation time or temperature, as well as kit age can cause a change in signal.

_Not all interfering factors have been tested in the immunoassay, therefore the possibility of interference cannot be excluded.

COMPONENTS PROVIDED

DESCRIPTION	CODE	QUANTITY
Albumin Microplate - 96 well polystyrene microplate (12 strips of 8 wells) coated with monoclonal antibody against Human albumin.	383-06-01	1 plate
Albumin Standard – lot specific of human albumin for calibration in a buffered protein base with preservative; lyophilized.	383-06-02	1 vial
Detection Antibody – lot specific, 10-fold concentrated rabbit antibody against Human albumin.	383-06-03	1 vial
Positive Control – one vial of human albumin; lyophilized.	383-06-04	1 vial
Anti Rabbit IgG-HRP Conjugate - 120 µL of 100-fold concentrated Anti IgG-HRP Conjugate.	ARIGHRP	1 vial
PBS-20x Concentrate - 25 mL of 20-fold concentrate PBS solution with preservative.	PBS-20X	1 Bottle
10X Dilution Buffer Concentrate - 40 mL of 10-fold concentrate buffered protein based solution with preservative.	10XDB08K	1 bottle
Antibody Diluent Solution - 30 mL of 10-fold concentrate buffered protein based solution with preservative.	DB08A	1 bottle
Wash Buffer - 50 mL of 10-fold concentrated buffered surfactant, with preservative.	WB02	1 bottle
TMB Substrate Solution - 11 mL of TMB substrate solution	TMB01	1 bottle
Stop Solution - 11 mL of 0.5M HCl	S-STOP	1 bottle
Plate Sealer	EAPS	1
Plastic Pouch	P01	1

STORAGE

Unopened Kit: Store at 2 - 8°C for up to 10 months. Anti Rabbit Antibody-HRP conjugate should be stored at 2°C ~ 8°C. Do not use kit past expiration date.

ADDITIONAL MATERIALS REQUIRED

- Microplate reader capable of absorbance measurement at 450 nm.
- Microplate shaker (250 – 300 rpm).
- Microplate washer or manifold dispenser.
- 100 mL and 500 mL graduated cylinders.
- Multi-channel Pipette, Pipettes and pipette tips.
- Deionized or distilled water.

PRECAUTION

This kit should be handled by those persons who have been trained in and can follow the principles of good laboratory practice. Wear protective clothing such as laboratory overalls, safety glasses and gloves. Care should be taken while handling solutions in this kit to avoid contact with skin or eyes, especially with the stop solution because it contains diluted hydrochloric acid. Wash immediately with water in case of contact on skin or eyes. Do not use any solutions contains bovine serum albumin in this ELISA assay.

SAMPLE COLLECTION AND STORAGE

Plasma - Collect plasma using EDTA, heparin, or citrate as an anticoagulant. Centrifuge for 15 minutes at 1000 x g within 30 minutes of collection. Assay immediately or aliquot and store samples at $\leq -20^{\circ}\text{C}$. Avoid repeated freeze-thaw cycles.

Serum - Use a serum separator tube (SST) and allow samples to clot for 30 minutes before centrifugation for 15 minutes at 1000 x g. Remove serum and assay immediately or aliquot and store samples at $\leq -20^{\circ}\text{C}$. Avoid repeated freeze-thaw cycles.

SAMPLE PREPARATION

Human serum and plasma samples may need an 1,000,000(1000K) ~ 4,000,000 (4000K)-fold dilution. A 100-fold dilution is 5 μL sample + 495 μL 1x Dilution Buffer. To make a 10,000-fold dilution is 5 μL of 100-fold sample + 495 μL 1x Dilution Buffer. Finally, to make a 1,000,000-fold dilution is 5 μL of 10,000-fold sample + 495 μL 1x Dilution Buffer. Finally, to make a 2,000,000-fold dilution is 120 μL of 1,000,000-fold sample + 120 μL 1x Dilution Buffer.

Finally, to make a 4,000,000-fold dilution is 80 μL of 1,000,000-fold sample + 240 μL 1x Dilution Buffer.

Optimal dilutions should be determined by each laboratory for each application.

Use polypropylene test tubes.

REAGENT PREPARATION

Bring all reagents to room temperature before use.

Wash Buffer - If crystals have formed in the concentrate, warm to room temperature and mix gently until the crystals have completely dissolved. Dilute 50 mL of Wash Buffer Concentrate into deionized or distilled water (450 mL) to prepare 500 mL of 1x Wash Buffer.

PBS-20X - If crystals have formed in the concentrate, warm to room temperature and mix gently until the crystals have completely dissolved. Dilute 25 mL of PBS-20X Concentrate into deionized or distilled water (475 mL) to prepare 500 mL of 1x PBS solution.

Dilution Buffer Concentrate (10XDB08K) - *10XDB08K cannot use directly. Must follow the dilution below:*

Dilute 40 mL of Dilution Buffer Concentrate (10-fold) into 360 mL of 1 x PBS solution to prepare 400 mL of 1x Dilution Buffer (DB08K).

Human Albumin Standard - Reconstitute the Albumin standard with lot specific of 1x Dilution Buffer. Pipette 250 μL of 1x Dilution Buffer into tubes #2 to #7. Use the stock solution to produce a dilution series (below). Mix each tube thoroughly before the next transfer. The **100 ng/mL** standard serves as the high standard. The Dilution Buffer serves as the zero standard (0 ng/mL).

Tube	Standard	1x Dilution Buffer	Concentration
Stock	Powder	Lot specific	
# 1	Lot specific	Lot specific	100 ng/ml
# 2	250 µl of 1	250 µl	50 ng/ml
# 3	250 µl of 2	250 µl	25 ng/ml
# 4	250 µl of 3	250 µl	12.5 ng/ml
# 5	250 µl of 4	250 µl	6.25 ng/ml
# 6	250 µl of 5	250 µl	3.125 ng/ml
# 7	250 µl of 6	250 µl	1.56 ng/ml

Positive Control - Reconstitute the positive control with lot specific of 1x Dilution Buffer to make positive control working solution.

Detection Antibody– Reconstitute the Detection Antibody with lot specific of Antibody Diluent Solution (DB08A) to make 10-fold **concentrate detection antibody solution**. Pipette lot specific of Antibody Diluent Solution (DB08A) into a 15 mL centrifuge tube and transfer lot specific of 10-fold concentrated stock solution to prepare working solution.

Anti Rabbit IgG-HRP Conjugate – Pipette 11.88 mL of **Antibody Diluent Solution (DB08A)** into a 15 mL centrifuge tube and transfer 120 µL of 100-fold concentrated stock solution to prepare working solution (**protect from light**).

ELISA PROTOCOL

Bring all reagents and samples to room temperature before the start of the assay. Blank, standard dilutions, positive control and samples should be assayed in duplicate. ELISA Protocol may need further optimization.

1. Prepare all reagents and working standards as directed in the previous sections.
2. Remove excess microplate strips from the plate frame, return them to the plastic pouch with the desiccant pack.
3. Add 100 µL per well of **1x Dilution Buffer** to Blank wells.
4. Add 100 µL of **Standard dilutions, samples, or positive control** per well. Cover with plate sealer.

Incubate for 2 hours on microplate shaker at room temperature.

5. Aspirate each well and wash, repeating the process three times for a total of four washes. Wash by filling each well with **1x Wash Buffer** (300 µL) using a squirt bottle, manifold dispenser, or autowasher. Complete removal of liquid at each step is essential to good performance. After the last wash, remove any remaining Wash Buffer by aspirating or decanting. Invert the plate and blot it against clean paper towels.
6. Add 100 µL of **Detection Antibody-HRP Conjugate working solution** to each well. Cover with plate sealer. Incubate for 1 hour on microplate shaker at room temperature. **Protect from light.**
7. Repeat the aspiration/wash as in step 5.
8. Add 100 µL of **Substrate Solution** to each well. Incubate for lot specific minutes. **Protect from light.** There may be fast color development, please be prepared to add stop solution immediately.
9. Add 100 µL of **Stop Solution** to each well. The color in the wells should change from blue to yellow. If the color in the wells is green, or if the color change does not appear uniform, gently tap the plate to ensure thorough mixing.
12. Determine the optical density of each well using a microplate reader set to 450 nm.

CALCULATION OF RESULTS

Average the duplicate readings for each standard, positive control and sample, and subtract the average zero standard optical density. Create a standard curve by reducing the data using computer software capable of generating a 4-parameter logistic (4-PL) curve fit.

If samples have been diluted, the concentration read from the standard curve must be multiplied by the dilution factor.

SPECIFICITY

PROTEINS	CROSS-REACTIVITY (%)
Human Serum Albumin	100
Human CRP	0
Human Transferrin	0
Human Fetuin A	0
Human Adiponectin	0
Human RBP-4	0

The serum samples from following species showed no significant cross-reactivity at 1:20000 dilution: mouse and rat.

TYPICAL DATA

The standard curve is provided for demonstration only. A new standard curve should be generated for each set of samples assayed.

STANDARD (NG/ML)	AVERAGE OD450NM
Blank	0 (lot specific)
1.56	0.035
3.125	0.076
6.25	0.132
12.5	0.288
25	0.658
50	1.162
100	2.008

Use 5 µL of Human serum or plasma samples to prepare 1: 1000K or 4000K dilution.

		Final Dilution
5µL of Human sample	495 µL of 1x Dilution Buffer (DB08K)	100
5µL of 100-fold diluted sample solution	495 µL of 1x Dilution Buffer (DB08K)	10000
5µL of 10000-fold diluted sample solution	495 µL of 1x Dilution Buffer (DB08K)	1000000 (1000K)
120 µL of 1000000-fold diluted sample solution	120 µL of 1x Dilution Buffer (DB08K)	2000000 (2000K)
80 µL of 1000000-fold diluted sample solution	240 µL of 1x Dilution Buffer (DB08K)	4000000 (4000K)

SUMMARY OF ASSAY PROCEDURE

PREPARE REAGENTS, SAMPLES AND STANDARDS
↓
Add 100 µl of standard dilutions, samples, or positive control to the well. Incubate 2 hours on the plate shaker at RT.
↓
Aspirate and wash 4 times.
↓
Add 100 µl of Detection Antibody working solution to each well. Incubate 2 hours on the plate shaker at RT.
↓
Aspirate and wash 4 times.
↓
Add 100 µl of Anti Rabbit IgG-HRP working solution to each well. Incubate 60 minutes on the plate shaker at RT. Protect from light.
↓
Aspirate and wash 4 times.
↓
Add 100 µl of TMB Substrate Solution to each well. Incubate lot specific min on the plate shaker at RT. Protect from light.
↓
Add 100 µl of Stop Solution to each well. Read at 450nm.

Use 10 µL of Human serum or plasma samples to prepare 1: 1000K or 4000K dilution.

		Final Dilution
10 µL of Human sample	995 µL of 1x Dilution Buffer (DB08K)	100
10 µL of 100-fold diluted sample solution	995 µL of 1x Dilution Buffer (DB08K)	10000
10 µL of 10000-fold diluted sample solution	995 µL of 1x Dilution Buffer (DB08K)	1000000 (1000K)
120 µL of 1000000-fold (1000K) diluted sample solution	120 µL of 1x Dilution Buffer (DB08K)	2000000 (2000K)
80 µL of 1000000-fold (1000K) diluted sample solution	240 µL of 1x Dilution Buffer (DB08K)	4000000 (4000K)