

## HUMAN IL-12 ELISA KIT

**FOR THE QUANTITATIVE DETERMINATION OF HUMAN IL-12 CONCENTRATIONS IN CELL CULTURE SUPERNATES, PLASMA AND SERUM**



**FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.**

### PURCHASE INFORMATION:

ELISA Name	Human IL-12 ELISA
Catalog No.	SK00295-02
Lot No.	
Formulation	96 T
Standard Range	4 – 250 pg/mL
Sensitivity	4 pg/mL
Sample Volume	100 µl
Sample Type	Serum, EDTA Plasma, Cell Culture
Dilution Factor	Optimal dilutions should be determined by each laboratory for each application
Specificity	Human IL-12 only
Intra-assay Precision	4-6%
Inter-assay Precision	8-10%
Storage	2°C - 8°C

### Order Contact:

**AVISCERA BIOSCIENCE INC.**

**2348 Walsh Ave., Suite C**

**Santa Clara, CA 95051**

**Tel: (408) 982 0300**

**Fax: (408) 982 0301**

**Email: [Sales@AvisceraBioscience.com](mailto:Sales@AvisceraBioscience.com)**

**[Info@AvisceraBioscience.com](mailto:Info@AvisceraBioscience.com)**

**[www.AvisceraBioscience.com](http://www.AvisceraBioscience.com)**

## INTRODUCTION

Human IL-12 immunoassay is a 3.5 - 4.5 hour solid phase ELISA designed to measure human IL-12 in cell culture supernates, serum and plasma. It contains recombinant human IL-12 and antibodies raised against this protein. It has been shown to accurately quantify recombinant human IL-12. Results obtained with naturally occurring IL-12 samples show linear curves that were parallel to the standard curves obtained using the kit standards. These results indicate that the immunoassay kit can be used to determine relative mass values for natural human IL-12.

## PRINCIPLE OF THE ASSAY

This assay employs the quantitative sandwich enzyme immunoassay technique. A monoclonal antibody specific for IL-12 has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and any IL-12 present is bound by the immobilized antibody. After washing away any unbound substances, a biotinylated monoclonal antibody specific for IL-12 is added to the wells. Following a wash to remove any unbound antibody-biotin reagent, HRP link Streptavidin is added to the wells. After washing away any unbound enzyme, a substrate solution is added to the wells and color develops in proportion to the amount of IL-12 bound in the initial step. The color development is stopped and the intensity of the color is measured.

## LIMITATIONS OF THE PROCEDURE

- \_ FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.
- \_ The kit should not be used beyond the expiration date on the kit label.
- \_ Do not mix or substitute reagents with those from other lots or sources.
- \_ It is important that the dilution buffer selected for the standard curve be consistent with the samples being assayed.
- \_ If samples generate values higher than the highest standard, dilute the samples with Dilution Buffer and repeat the assay.
- \_ Any variation in standard diluent, operator, pipetting technique, washing technique, incubation time or temperature, and kit age can cause variation in binding.
- \_ This assay is designed to eliminate interference by soluble receptors, binding proteins, and other factors present in biological samples. Until all factors

have been tested in the immunoassay, the possibility of interference cannot be excluded.

## MATERIALS PROVIDED

Description	Code	Quantity
<b>Microplate</b> - 96 well polystyrene microplate (12 strips of 8 wells) coated with a monoclonal antibody against IL-12.	295-02-01	1 plate
<b>IL-12 Standard</b> – 1000 pg/vial of recombinant human IL-12 in a buffered protein base with preservatives; lyophilized.	295-02-02	1 vial
<b>Detection Antibody Concentrate</b> – 120 µL / vial, 100-fold concentrated of Biotinylated monoclonal antibody against IL-12 with preservatives; lyophilized.	295-02-03	1 vial
<b>Positive Control</b> - one vial of recombinant human IL-12, lyophilized	295-02-04	1 vial
<b>Streptavidin-HRP Conjugate</b> - 120 ul/vial, 100-fold concentrated solution of Streptavidin conjugate to HRP	SAHRP	1 vial
<b>Dilution Buffer</b> - 60mL/vial of buffered protein based solution with preservatives	DB07	1 vial
<b>HRP Dilution Solution</b> – 12mL/vial of buffered protein based solution with preservatives	DB06C	1 vial
<b>Wash Buffer</b> - 50 ml/vial, 10-fold concentrated buffered surfactant, with preservative.	WB01	1 vial
<b>TMB Substrate Solution</b> - 11ml/vial of TMB substrate solution	TMB01	1 vial
<b>Stop Solution (0.5M HCl)</b> , 11ml/vial of 0.5M HCl	S-STOP	1 vial
<b>Plate Covers</b> – Plate sealer	EAPS	1

**STORAGE**

**Unopened Kit:** Store at 2 - 8°C for up to 6 months. For longer storage, unopened Standard, Positive Control and Detection Antibody Concentrate as well as Dilution Buffer (DB07) should be stored at -20°C or -70°C. Do not use kit past expiration date.

**Opened / Reconstituted Reagents:** Reconstituted Standard , Antibody Solution SHOULD BE STORED at -20°C or - 70°C for up to one month. Streptavidin-HRP Conjugate 100-fold concentrate and other components may be stored at 2 - 8°C for up to 6 months.

**Microplate Wells:** Return unused wells to the plastic bag containing the desiccant pack, reseal along entire edge of zip-seal. May be stored for up to 6 months at 2 - 8°C.

**OTHER SUPPLIES REQUIRED**

- Microplate reader capable of measuring absorbance at 450 nm, with the correction wavelength set at 540 nm or 570 nm.
- Microplate shaker (250-300rpm).
- Pipettes and pipette tips.
- Deionized or distilled water.
- Squirt bottle, manifold dispenser, or automated microplate washer.
- 100 mL and 500 mL graduated cylinders.

**PRECAUTIONS FOR USE**

All reagents should be considered as potentially hazardous. The stop solution contains diluted Hydrochloric acid. Appropriate care should be taken while handling this solution. We recommend that this product be handled by persons who have been trained in laboratory techniques and that it is used in accordance with the principles of good laboratory practice. Wear suitable protective clothing such as laboratory overalls, safety glasses and gloves. Care should be taken to avoid contact with skin or eyes. In the case of contact with skin or eyes wash immediately with water.

**SAMPLE COLLECTION AND STORAGE**

**Plasma** - Collect plasma using EDTA, heparin, or citrate as an anticoagulant. Centrifuge for 15 minutes at 1000 x g within 30 minutes of collection. Assay immediately or aliquot and store samples at ≤-20°C. Avoid repeated freeze-thaw cycles. Optimal dilutions should be determined by each laboratory for each application.


**REAGENT PREPARATION**

**Bring all reagents to room temperature before use.**

**Wash Buffer** - If crystals have formed in the concentrate, warm to room temperature and mix gently until the crystals have completely dissolved. Dilute 50 mL of Wash Buffer Concentrate into deionized or distilled water (450 mL) to prepare 500 mL of Wash Buffer.

**IL-12 Standard - Refer to vial label for reconstitution volume.** Reconstitute the **IL-12** Standard with 1 ml of Dilution Buffer. This reconstitution produces a stock solution of 1000 pg/mL. Allow the standard to sit for a minimum of 15 minutes with gentle agitation prior to making dilutions. Pipette 600µL of Dilution Buffer into tube #1, then transfer 200µL of 1000 pg/mL stock solution to tube #1. This will create the high standard of 250 pg/mL). Then, pipette 250µL of Dilution Buffer into tubes #2 to #7. Use the high standard solution to produce a dilution series (see below). Mix each tube thoroughly before the next transfer. The 250 pg/mL standard serves as the high standard. The Dilution Buffer serves as the zero standard (0 pg/mL).

Standard	Standard	Reagent Diluent	Concentration
Stock	Powder	1000 µl	1000 pg/mL
# 1	200 µl of stock	600 µl	250 pg/mL
# 2	250 µl of 1	250 µl	125 pg/mL
# 3	250 µl of 2	250 µl	62.5 pg/mL
# 4	250 µl of 3	250 µl	31.3 pg/mL
# 5	250 µl of 4	250 µl	15.6 pg/mL
# 6	250 µl of 5	250 µl	7.81 pg/mL
#7	250 µl of 6	250 µl	3.91 pg/mL



**Positive Control** - Reconstitute the **Positive Control** with 1.0 mL Dilution Buffer. *Positive Control should be prepared and used immediately.*

**Detection Antibody-** Reconstitute the **Detection Antibody Concentrate** with 120  $\mu\text{L}$  of Dilution Buffer to produce a 100-fold concentrated stock solution. Pipette 11.88 mL of Dilution Buffer into a 15 ml centrifuge tube and transfer 120  $\mu\text{L}$  of 100-fold concentrated stock solution to prepare working solution.

**Streptavidin-HRP Conjugate** - Pipette 11.88 mL of HRP Dilution Solution (DB06C) into a 15 ml centrifuge tube and transfer 120  $\mu\text{L}$  of the 100-fold concentrated stock solution to prepare working solution.

### ASSAY PROCEDURE

**Bring all reagents and samples to room temperature before use. It is recommended that standards be assayed in duplicates.**


1. Prepare all reagents and working standards as directed in the previous sections.
2. Remove excess micro-plate strips from the plate frame, return them to the foil pouch containing the desiccant pack, reseal.
3. Add 100  $\mu\text{L}$  of Dilution Buffer to Blank well (A2,A3).
4. Add 100  $\mu\text{L}$  of Standard (from B2,B3 to G2,G3 and G4,G5), sample, or control (F4,F5) per well. Cover with plate sealer. Incubate for 2 hours on micro-plate shaker at room temperature. A plate layout is provided to record standards and samples assayed.
5. Aspirate each well and wash, repeating the process three times for a total of four washes. Wash by filling each well with Wash Buffer (300  $\mu\text{L}$ ) using a squirt bottle, manifold dispenser, or autowasher. Complete removal of liquid at each step is essential to good performance. After the last wash, remove any remaining Wash Buffer by aspirating or decanting. Invert the plate and blot it against clean paper towels.
6. Add 100  $\mu\text{L}$  of Detection Antibody working solution to each well. Cover with plate sealer. Incubate for 2 hours on micro-plate shaker at room temperature.
7. Repeat the aspiration/wash as in step 5.
8. Add 100  $\mu\text{L}$  of **Streptavidin-HRP Conjugate** working solution to each well. Incubate for 45 minutes on micro-plate shaker at room temperature. **Protect from light.**
9. Repeat the aspiration/wash as in step 5.

10. Add 100  $\mu\text{L}$  of Substrate Solution to each well. Incubate for 5-15 minutes at room temperature. **Protect from light.**
11. Add 100  $\mu\text{L}$  of Stop Solution to each well. The color in the wells should change from blue to yellow. If the color in the wells is green or if the color change does not appear uniform, gently tap the plate to ensure thorough mixing.
12. Determine the optical density of each well within 15 minutes, using a micro-plate reader set to 450 nm.

### CALCULATION OF RESULTS

Average the duplicate readings for each standard, control, and sample and subtract the average zero standard optical density. Create a standard curve by reducing the data using computer software capable of generating a log-log curve fit. As an alternative, construct a standard curve by plotting the mean absorbance for each standard on the y-axis against the concentration on the x-axis and draw a best fit curve through the points on the graph. The data may be linearized by plotting the log of the IL-12 concentrations versus the log of the O.D. and the best fit line can be determined by regression analysis. This procedure will produce an adequate but less precise fit of the data.

If samples have been diluted, the concentration read from the standard curve must be multiplied by the dilution factor.



### TYPICAL DATA

These standard curves\* are provided for demonstration only. A standard curve should be generated for each set of samples assayed.

IL-12(pg/ml)	Average OD450 (Corrected)*
3.906	0.064
7.813	0.101
15.625	0.186
31.25	0.347
62.5	0.745
125	1.564
250	1.954

\*Lot No.:

\*\* Positive Control: 97 - 162 pg/ml

**CALIBRATION**

This immunoassay is calibrated against a highly purified E. Coli-expressed recombinant human IL-12.

**SENSITIVITY**

Twenty-five assays were evaluated and the minimum detectable dose (MDD) of IL-12 was 4 pg/mL.

**SPECIFICITY**

This assay recognizes both natural and recombinant human IL-12. The factors listed below were prepared at 50 ng/mL in Dilution Buffer, and assayed for cross reactivity. Preparations of the following factors at 50 ng/mL in a mid-range rh IL-12 control were assayed for interference. No significant cross-reactivity or interference was observed.

Human Recombinant Proteins:

IL-10, IL-6, IL-8, IL-2, IL-1beta

Mouse Recombinant Proteins:

IL-12

**SUMMARY OF ASSAY PROCEDURE**

