

Human SOLUBLE CD305 ELISA KIT

FOR THE QUANTITATIVE DETERMINATION
OF HUMAN SOLUBLE CD305
CONCENTRATIONS IN SERUM AND CELL
CULTURES.



THIS IS DEMONSTRATION ONLY.

FOR RESEARCH USE ONLY. NOT FOR USE
IN DIAGNOSTIC PROCEDURES.

PURCHASE INFORMATION:

ELISA NAME	HUMAN SOLUBLE CD305 ELISA
Catalog No.	SK00410-01
Lot No.:	
Formulation	96 T
Standard range	93 - 6000 pg/ml
Sensitivity	50 pg/mL
Sample Volume	100 µl
Dilution Factor	<i>Optimal dilutions should be determined by each laboratory for each application</i>
Sample Type	Serum, cell culture
Specificity	Human Soluble CD305
Intra-assay Precision	4-6%
Inter-assay Precision	8-12%
Storage	4 °C

ORDER CONTACT:

AVISCIERA BIOSCIENCE INC.
2348 Walsh Ave., Suite C
Santa Clara, CA 95051
Tel: (408) 982 0300
Fax: (408) 982 0301
Email: Sales@AvisceraBioscience.com
Info@AvisceraBioscience.com
www.AvisceraBioscience.com

INTRODUCTION

Human soluble CD305 ELISA kit is a solid phase ELISA designed to measure human soluble CD305 in cell culture supernates and serum. It contains recombinant human soluble CD305 and antibodies raised against this protein. It has been shown to accurately quantitate recombinant human soluble CD305. Results obtained with naturally occurring soluble CD305 samples showed linear curves that were parallel to the standard curves obtained using the kit standards. These results indicate that the Immunoassay kit can be used to determine relative mass values for natural soluble CD305.

PRINCIPLE OF THE ASSAY

This assay employs the quantitative sandwich enzyme immunoassay technique. A monoclonal antibody specific for human soluble CD305 has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and any CD305 present is bound by the immobilized antibody. After washing away any unbound substances, a polyclonal antibody specific for CD305 is added to the wells. Following a wash to remove any unbound antibody reagent, Anti rabbit IgG HRP link Streptavidin is added to the wells. After washing away any unbound enzyme, a substrate solution is added to the wells and color develops in proportion to the amount of CD305 bound in the initial step. The color development is stopped and the intensity of the color is measured.

LIMITATIONS OF THE PROCEDURE

_ FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.

_ The kit should not be used beyond the expiration date on the kit label.

_ Do not mix or substitute reagents with those from other lots or sources.

_ It is important that the Dilution Buffer selected for the standard curve be consistent with the samples being assayed.

_ If samples generate values higher than the highest standard, dilute the samples with the appropriate Dilution Buffer and repeat the assay.

_ Any variation in standard diluent, operator, pipetting technique, washing technique, incubation time or temperature, and kit age can cause variation in binding.

_ This assay is designed to eliminate interference by soluble receptors, binding proteins, and other factors present in biological samples. Until all factors

have been tested in the Immunoassay, the possibility of interference cannot be excluded.

PRECAUTIONS FOR USE

All reagents should be considered as potentially hazardous. The stop solution contains diluted Hydrochloric acid. Appropriate care, therefore, should be taken while handling this solution. We therefore recommend that this product is handled only by those persons who have been trained in laboratory techniques and that it is used in accordance with the principles of good laboratory practice. Wear suitable protective clothing such as laboratory overalls, safety glasses and gloves. Care should be taken to avoid contact with skin or eyes. In the case of contact with skin or eyes wash immediately with water.

MATERIALS PROVIDED

DESCRIPTION	CODE	QUANTITY
CD305-Microplate – 96 well microplate precoated with monoclonal anti-human soluble CD305, one plate	410-01-01	1 plate
CD305 Standard – refer to lot specific of recombinant soluble CD305 in a buffered protein base with preservatives; lyophilized.	410-01-02	1 vial
CD305 Antibody Concentrate —refer to lot specific concentrated of polyclonal Antibody against human soluble CD305 with preservatives; lyophilized.	410-01-03	1 vial
Positive Control – one vial of recombinant human soluble CD305 , lyophilized (optional)	410-01-04	1 vial
Anti Rabbit IgG -HRP Conjugate -120 µl/vial, 100-fold concentrated solution of Streptavidin conjugate to HRP	ARIGHRP	1 vial
Dilution Buffer - 60 mL/vial of buffered protein based solution with preservatives	DB08B	1 vial
Wash Buffer -50 ml/vial, 10-fold concentrated buffered	WB01	1 vial

surfactant, with preservative.		
TMB Substrate Solution- 11ml / vial of TMB substrate solution	TMB01	1 vial
Stop Solution (0.5M HCl) , 11 ml /vial of 0.5M HCl	S-STOP	1 vial
Plate Sealer – Plate sealer.	EAPS	1

STORAGE

Unopened Kit: Store at 2 - 8° C for up to 6 months. For longer storage, unopened Standard, Positive Control and Antibody Concentrated should be stored at -20 or -70 °C. Do not use past kit expiration date. Streptavidin - HRP Conjugate 100-fold concentrated and other components may be stored at 2 - 8°C for up to 6 months.

Microplate Wells: Return unused wells to the plastic bag containing the desiccant pack, reseal along entire edge of zip-seal. May be stored for up to 4 months at 2 - 8° C.

OTHER SUPPLIES REQUIRED

- Microplate reader capable of measuring absorbance at 450 nm, with the correction wavelength set at 540 nm or 570 nm.
- Microplate shaker (250-300rpm).
- Pipettes and pipette tips.
- Deionized or distilled water.
- Squirt bottle, manifold dispenser, or automated microplate washer.
- 100 mL and 500 mL graduated cylinders.

SAMPLE COLLECTION AND STORAGE

Cell Culture Supernates - Remove particulates by centrifugation and assay immediately or aliquot and store samples at ≤-20° C. Avoid repeated freeze-thaw cycles.

Serum - Use a serum separator tube (SST) and allow samples to clot for 30 minutes before centrifugation for 15 minutes at 1000 x g. Remove serum and assay immediately or aliquot and store samples at ≤ -20° C. Avoid repeated freeze-thaw cycles.

SAMPLE PREPARATION

Optimal dilutions should be determined by each laboratory for each application.

Use polypropylene test tubes.

REAGENT PREPARATION

Bring all reagents to room temperature before use.

Wash Buffer - If crystals have formed in the concentrate, warm to room temperature and mix gently until the crystals have completely dissolved. Dilute 50 mL of Wash Buffer Concentrate into deionized or distilled water (450 mL) to prepare 500 mL of Wash Buffer.

CD305 Standard - Refer to vial label for reconstitution volume. Reconstitute the **CD305** Standard with refer to lot specific of Dilution Buffer. Pipette 250 µL of the appropriate Dilution Buffer into the tube #2 to #7. Use the stock solution to produce a dilution series (below). Mix each tube thoroughly before the next transfer. The 5000 pg/mL standard serves as the high standard. The appropriate Dilution Buffer DB18 serves as the zero standard (0 pg/mL).

STANDARD TUBE	STANDARD	DILUTION BUFFER	CONCENTRATION
stock	powder	Refer to lot specific	
# 1	Refer to lot specific	Refer to lot specific	6000 pg/ml
# 2	250µl of 1	250µl	3000 pg/ml
# 3	250µl of 2	250µl	1500 pg/ml
# 4	250µl of 3	250µl	750 pg/ml
# 5	250µl of 4	250µl	375 pg/ml
# 6	250µl of 5	250µl	187.5 pg/ml
# 7	250µl of 6	250µl	93.75 pg/ml

CD305 Antibody- Reconstitute the **Antibody concentrated** with refer to lot specific of Dilution Buffer to produce a concentrated stock solution. Transfer it to refer to lot specific mL of Dilution Buffer to prepare 1 x Antibody solution.

Anti Rabbit IgG-HRP Conjugate - Transfer 120 µl of 100-fold concentrated stock solution to 12 ml of HRP Diluent Solution to prepare working solution. Note: 1 x working solution of Anti Rabbit IgG HRP Conjugate should be used within a few days.

Positive Control- Reconstitute the **Positive Control** with **refer to lot specific** of Dilution Buffer. Positive Control should be prepared and used immediately.

ASSAY PROCEDURE

Bring all reagents and samples to room temperature before use. It is recommended that standards be assayed in duplicate.

1. Prepare all reagents and working standards as directed in the previous sections.
2. Remove excess micro-plate strips from the plate frame, return them to the foil pouch containing the desiccant pack, reseal.
3. Leave well A2 and A3 as Blank. Add 100 µl per well of Dilution Buffer.
4. Add 100 µl per well of standard solution from #7 to #1 (reverse order of serial dilution) to the appropriate wells (B2 to G3, G4 to F5,). Add 100 µl per well of Positive control into well E4 and E5. Add 100 µl per well of samples into appropriate wells. Cover or seal the plate and incubate at room temperature for 2 hours on microplate shaker (250 rpm). Note: Standard, Blank and PC should be assayed in duplicate.
5. Aspirate wells and wash 4 times with 300 µl of 1 x Assay Wash Buffer. Blot plate on absorbent paper to remove any residual buffer.
6. Add 100 µl per well of 1 x Antibody solution. Cover or seal the plate and incubate at room temperature for 1 hour on microplate shaker (250 rpm).
7. Repeat the aspiration/wash as in step 5.
8. Add 100 µL of **Anti Rabbit IgG-HRP Conjugate** working solution. Cover or seal the plate and incubate at room temperature for 30 minutes on microplate shaker.
11. Repeat the aspiration/wash as in step 5.
12. Add 100 µL of Substrate Solution to each well. Incubate for refer to lot specific at room temperature. **Protect from light.**
13. Add 100 µL of Stop Solution to each well. The color in the wells should change from blue to yellow. If the color in the wells is green or if the color change does not appear uniform, gently tap the plate to ensure thorough mixing.
12. Determine the optical density of each well within 15 minutes, using a micro-plate reader set to 450 nm.

CALCULATION OF RESULTS

Average the duplicate readings for each standard, QC, and samples and subtract the average Blank optical density. It is recommended to use software capable of generating a log-log curve-fit. The standard curve shows relationship between standard concentrations and corresponding O.D absorbance. If samples have been diluted, the concentration read from the standard curve must be multiplied by the dilution factor.

CALIBRATION

This immunoassay is calibrated against a highly purified recombinant human soluble CD305 .

SENSITIVITY

The minimum detectable dose (MDD) of human soluble CD305 was 50 pg/mL.

TYPICAL DATA








These standard curves are provided for demonstration only. A standard curve should be generated for each set of samples assayed.

STANDARD (PG/ML)	OD450 READING
0 (Blank)	0 (0.065)
93.75	0.059
187.5	0.111
375	0.209
750	0.401
1500	0.736
3000	1.127
6000	1.811

SPECIFICITY

PROTEIN	CROSSREACTIVITY (%)
Human Soluble CD305	100
Human Soluble CD306	0
Human Soluble CD209	0
Human Soluble CD36	0
Human Soluble CD10	0

SUMMARY OF ASSAY PROCEDURE

PREPARE REAGENTS, SAMPLES AND STANDARDS

Add 100µl of standard, samples, positive control to each well. Incubate 2 hours on the plate shaker at RT.

Aspirate and wash 4 times.

Add 100 µl Antibody Solution to each well. Incubate 1 hour on the plate shaker at RT.

Add 100 µl Anti Rabbit IgG HRP conjugate to all wells. Incubate 30 min on the plate shaker at RT.

Aspirate and wash 4 times.

Add 100 µl Substrate Solution to each well. Incubate REFER TO LOT SPECIFIC on the bench top. Protect from light.

Add 100 µl Stop Solution to each well. Read 450nm within 15 min