

HUMAN SOLUBLE LAIR2/CD302 ELISA KIT

FOR THE QUANTITATIVE DETERMINATION
OF SOLUBLE LAIR2/CD306
CONCENTRATIONS IN HUMAN SERUM AND
PLASMA



FOR RESEARCH USE ONLY. NOT FOR USE IN
DIAGNOSTIC PROCEDURES.

PURCHASE INFORMATION:

ELISA NAME	HUMAN SOLUBLE LAIR2/CD306 ELISA
Catalog No.	SK00411-01
Formulation	96 T
Lot No.	
Standard range	470-30000 pg/ml
Sensitivity	80 pg/ml
Sample Volume	100 µl
Dilution	Optimal dilutions should be determined by each laboratory for each application
Sample Type	serum, plasma
Specificity	Human LAIR2/CD306 only
Intra-assay Precision	4-8%
Inter-assay Precision	6-8%
Storage	2°C-8°C

Order Contact:

AVISCIERA BIOSCIENCE INC.

2348 Walsh Ave., Suite C

Santa Clara, CA 95051

Tel: (408) 982 0300

Fax: (408) 982 0301

Email: Info@AvisceraBioscience.com

www.AvisceraBioscience.com

INTRODUCTION

Human Soluble LAIR2/CD306 immunoassay is a 3.5 - 4.5 hour solid phase ELISA designed to measure soluble LAIR2/CD306 in human serum and plasma. It contains recombinant soluble LAIR2/CD306 and antibodies raised against this protein. It has been shown to accurately quantify recombinant LAIR2/CD306. Results obtained with naturally occurring soluble LAIR2/CD306 samples showed linear curves that were parallel to the standard curves obtained using the kit standards. These results indicate that the immunoassay kit can be used to determine relative mass values for natural soluble LAIR2/CD306.

PRINCIPLE OF THE ASSAY

This assay employs the quantitative sandwich enzyme immunoassay technique. A monoclonal antibody specific for soluble LAIR2/CD306 has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and any soluble LAIR2/CD306 present is bound by the immobilized antibody. After washing away any unbound substances, a biotinylated polyclonal antibody specific for soluble LAIR2/CD306 is added to the wells. Following a wash to remove any unbound antibody-biotin reagent, HRP link streptavidin is added to the wells. After washing away any unbound enzyme, a substrate solution is added to the wells and color develops in proportion to the amount of soluble LAIR2/CD306 bound in the initial step. The color development is stopped and the intensity of the color is measured.

LIMITATIONS OF THE PROCEDURE

_ FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.

_ The kit should not be used beyond the expiration date on the kit label.

_ Do not mix or substitute reagents with those from other lots or sources.

_ It is important that the Dilution Buffer selected for the standard curve be consistent with the samples being assayed.

_ If samples generate values higher than the highest standard, dilute the samples with Dilution Buffer and repeat the assay.

_ Any variation in standard diluent, operator, pipetting technique, washing technique, incubation time or temperature, and kit age can cause variation in binding.

_ This assay is designed to eliminate interference by soluble receptors, binding proteins, and other factors present in biological samples. Until all factors have been tested in the immunoassay, the possibility of interference cannot be excluded.

MATERIALS PROVIDED

DESCRIPTION	CODE	QUANTITY
Soluble LAIR2/CD306 Microplate - 96 well polystyrene microplate (12 strips of 8 wells) coated with a monoclonal antibody against LAIR2/CD306.	411-01-01	1 plate
LAIR2/CD306 Standard – 30000 pg/vial of recombinant soluble LAIR2/CD306 in a buffered protein base with preservatives; lyophilized.	411-01-02	1 vial
Detection Antibody Concentrate – 120 µL/vial, 100-fold concentrate of biotinylated polyclonal antibody against Soluble LAIR2/CD306 with preservatives; lyophilized.	411-01-03	1 vial
Positive Control - one vial of recombinant soluble LAIR2/CD306, lyophilized	411-01-04	1 vial
Streptavidin-HRP Conjugate - 75 ul/vial, 200-fold concentrated solution of Streptavidin conjugate to HRP with preservatives	SAHRP	1 vial
Dilution Buffer - 60ml of buffered protein based solution with preservatives	DB01	1 bottle
Wash Buffer - 50 ml of 10-fold concentrated buffered surfactant, with preservative.	WB01	1 bottle
TMB Substrate Solution - 11 ml of TMB substrate solution	TMB01	1 bottle
Stop Solution – 11ml of 0.5M HCl	S-STOP	1 bottle
Plate Sealer	EAPS	1 piece

STORAGE

Unopened Kit: Store at 2 - 8° C for up to 6 months. For longer storage, unopened Standard, Positive control and Detection Antibody Concentrate should be stored at -20 or -70°C. Do not use kit past expiration date.

Opened / Reconstituted Reagents: Reconstituted standard and detection antibody concentrate could be stored for up to one month at -70°C. Diluted standard working solution and positive control should be prepared and used immediately. Streptavidin-HRP Conjugate 200-fold concentrate and other components may be stored at 2 - 8°C for up to 6 months.

Microplate Wells: Return unused wells to the plastic pouch containing the desiccant pack, reseal along entire edge of zip-seal. Microplate may be stored for up to 6 months at 2 - 8°C.

OTHER SUPPLIES REQUIRED

- Microplate reader capable of measuring absorbance at 450 nm, with the correction wavelength set at 540 nm or 570 nm.
- Microplate shaker (250-300rpm).
- Pipettes and pipette tips.
- Deionized or distilled water.
- Squirt bottle, manifold dispenser, or automated microplate washer.
- 100 ml and 500 ml graduated cylinders.

SAMPLE COLLECTION AND STORAGE

Serum - Use a serum separator tube (SST) and allow samples to clot for 30 minutes before centrifugation for 15 minutes at 1000 x g. Remove serum and assay immediately or aliquot and store samples at ≤ -20° C. Avoid repeated freeze-thaw cycles.

Plasma - Collect plasma using EDTA, heparin, or citrate as an anticoagulant. Centrifuge for 15 minutes at 1000 x g within 30 minutes of collection. Assay immediately or aliquot and store samples at ≤ -20° C. Avoid repeated freeze-thaw cycles.

Note: Use Aprotinin (enzyme inhibitor) for ALL sample collection to prevent sample degradation. 0.5 TIU per ml of sample solution.

REAGENT PREPARATION


Bring all reagents to room temperature before use.

Wash Buffer - If crystals have formed in the concentrate, warm to room temperature and mix gently until the crystals have completely dissolved. Dilute 50 ml of Wash Buffer Concentrate into

deionized or distilled water (450 ml) to prepare 500 ml of Wash Buffer.

Soluble LAIR2/CD306 Standard - Refer to vial label for reconstitution volume. Reconstitute the soluble LAIR1/CD305 Standard with 1.0 mL of Dilution Buffer. This reconstitution produces a stock solution of 30000 pg/mL. Allow the standard to sit for a minimum of 15 minutes with gentle agitation prior to making dilutions. Pipette 250 µL of Dilution Buffer into tubes #1 to #6. Use the stock solution to produce a dilution series (below). Mix each tube thoroughly before the next transfer. The 30000 pg/mL standard serves as the high standard. The Dilution Buffer serves as the zero standard (0 pg/mL).

TUBE	STANDARD	DILUTION BUFFER	CONCENTRATION
stock	Powder	1000 µl	30000 pg/ml
# 1	250 µl of stock	250 µl	15000 pg/ml
# 2	250 µl of 1	250 µl	7500 pg/ml
# 3	250 µl of 2	250 µl	3750 pg/ml
# 4	250 µl of 3	250 µl	1875 pg/ml
# 5	250 µl of 4	250 µl	937.5 pg/ml
# 6	250 µl of 5	250 µl	468.75 pg/ml



Detection Antibody - Reconstitute the **Detection Antibody Concentrate** with 120 µL of Dilution Buffer to produce a 100-fold concentrated stock solution. Pipette 11.88 mL of Dilution Buffer into a 15 mL centrifuge tube and transfer 120 µl of 100-fold concentrated stock solution to prepare working solution.

Streptavidin-HRP Conjugate - Pipette 11.94 mL of Dilution Buffer into a 15 mL centrifuge tube and transfer 60 µL of 200-fold concentrated stock

solution to prepare working solution. **Note:** 1X working solution of Streptavidin-HRP conjugate should be used within a few days.

Positive Control - Reconstitute the **Positive Control** with 1.0 mL of Dilution Buffer. **Note:** Positive Control should be prepared and used immediately.


ASSAY PROCEDURE

Bring all reagents and samples to room temperature before use. It is recommended that standards and positive control be assayed in duplicates.

1. Prepare all reagents and working standards as directed in the previous sections.
2. Remove excess micro-plate strips from the plate frame, return them to the plastic pouch containing the desiccant pack, reseal.
3. Add 100 μ L of **Dilution Buffer** to Blank wells (F4, F5).
4. Add 100 μ L of **Standard** (B2, B3 to G2, G3 and G4, G5), **sample**, or **positive control** per well (E4, E5). Cover with plate sealer. Incubate for 2 hours on micro-plate shaker at room temperature. A plate layout is provided to record standards and samples assayed.
5. Aspirate each well and wash, repeating the process three times for a total of four washes. Wash by filling each well with **1X Wash Buffer** (300 μ L) using a squirt bottle, manifold dispenser, or autowasher. Complete removal of liquid at each step is essential to good performance. After the last wash, remove any remaining Wash Buffer by aspirating or decanting. Invert the plate and blot it against clean paper towels.
6. Add 100 μ L of **Detection Antibody working solution** to each well. Cover with sealer. Incubate for 2 hours on micro-plate shaker at room temperature.
7. Repeat the aspiration/wash as in step 5.
8. Add 100 μ L of **Streptavidin-HRP Conjugate working solution** to each well. Incubate for 60 minutes on micro-plate shaker at room temperature. **Protect from light.**
9. Repeat the aspiration/wash as in step 5.
10. Add 100 μ L of **Substrate Solution** to each well. Incubate for 3-7 minutes at room temperature. **Protect from light.**
11. Add 100 μ L of **Stop Solution** to each well. The color in the wells should change from blue to yellow. If the color in the wells is green, or if the

color change does not appear uniform, gently tap the plate to ensure thorough mixing.

12. Determine the optical density of each well within 15 minutes, using a micro-plate reader set to 450 nm.



CALCULATION OF RESULTS

Average the duplicate readings for each standard, positive control, and sample and subtract the average zero standard optical density. Create a standard curve by reducing the data using computer software capable of generating a log-log curve fit. As an alternative, construct a standard curve by plotting the mean absorbance for each standard on the y-axis against the concentration on the x-axis and draw a best fit curve through the points on the graph. The data may be linearized by plotting the log of the standard concentrations versus the log of the O.D. and the best fit line can be determined by regression analysis. This procedure will produce an adequate but less precise fit of the data.

If samples have been diluted, the concentration read from the standard curve must be multiplied by the dilution factor.

CALIBRATION

This immunoassay is calibrated against a highly purified recombinant Soluble LAIR2/CD306.

SENSITIVITY

Twenty-five assays were evaluated and the minimum detectable dose (MDD) of LAIR2/CD306 was 80 pg/mL.

TYPICAL DATA

This standard curve is provided for demonstration only. A standard curve should be generated for each set of samples assayed.

STANDARD (NG/ML)	AVERAGE OD450 (CORRECTED)
Blank	0 (0.137)
0.47	0.052
0.94	0.106
1.875	0.192
3.75	0.407
7.5	0.782
15	1.740
30	2.374


- Lot No.:
- Positive Control:

SPECIFICITY

This assay recognizes both natural and recombinant human LAIR2/CD306. The factors listed below were prepared at 1000 ng/ml in Dilution Buffer, and assayed for cross reactivity. No significant cross-reactivity or interference was observed.

Human Recombinant Proteins:
sCD36, sCD305

Mouse Recombinant Proteins:
sCD36



SUMMARY OF ASSAY PROCEDURE