

HUMAN SOLUBLE NEUREGULIN 1-β1 (NRG 1- β1) ELISA KIT

FOR THE QUANTITATIVE DETERMINATION
OF HUMAN NRG 1-β1 CONCENTRATIONS
IN SERUM AND EDTA PLASMA



PURCHASE INFORMATION:

ELISA Name	Human NRG 1-β1 ELISA
Catalog No.	SK00556-01
Lot No.:	
Formulation	96 T
Standard Range	125 - 8000 pg/mL
Sensitivity	62.5 pg/mL
Sample Volume	100 μL
Sample Type	Serum, EDTA Plasma
Specificity	Human NRG 1-β1
Sample Dilution	8 (<i>Optimal dilutions should be determined by each laboratory for each application.</i>)
Intra-assay Precision	4-6%
Inter-assay Precision	8-12%
Storage	2-8°C

ORDER CONTACT:

AVISCERA BIOSCIENCE INC.

2348 Walsh Ave., Suite C

Santa Clara, CA 95051

Tel: (408) 982 0300

Fax: (408) 982 0301

Email: Sales@AvisceraBioscience.com

Info@AvisceraBioscience.com

www.AvisceraBioscience.com

FOR RESEARCH USE ONLY. NOT FOR USE IN
DIAGNOSTIC PROCEDURES.

INTRODUCTION

Human NRG 1-β1 immunoassay is a 3.5 - 4.5 hour solid phase ELISA designed to measure human NRG 1-β1 in serum and plasma. It contains recombinant human NRG 1-β1 and antibodies raised against this protein. It has been shown to accurately quantify recombinant human NRG 1-β1. Results obtained with naturally occurring NRG 1-β1 samples showed linear curves that were parallel to the standard curves obtained using the kit standards. These results indicate that the immunoassay kit can be used to determine relative mass values for natural human NRG 1-β1.

PRINCIPLE OF THE ASSAY

This assay employs the quantitative sandwich enzyme immunoassay technique. A monoclonal antibody specific for NRG 1-β1 has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and any NRG 1-β1 present is bound by the immobilized antibody. After washing away any unbound substances, a biotinylated polyclonal antibody specific for NRG 1-β1 is added to the wells. Following a wash to remove any unbound antibody-biotin reagent, HRP link Streptavidin is added to the wells. After washing away any unbound enzyme, a substrate solution is added to the wells and color develops in proportion to the amount of NRG 1-β1 bound in the initial step. The color development is stopped and the intensity of the color is measured.

LIMITATIONS OF THE PROCEDURE

__ FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.

_ The kit should not be used beyond the expiration date on the kit label.

_ Do not mix or substitute reagents with those from other lots or sources.

_ It is important that the Dilution Buffer selected for the standard curve be consistent with the samples being assayed.

_ If samples generate values higher than the highest standard, dilute the samples with the appropriate Dilution Buffer and repeat the assay.

_ Any variation in standard diluent, operator, pipetting technique, washing technique, incubation time or temperature, and kit age can cause variation in binding.

_ Some vials contain small quantities of material, therefore centrifuge before use.

MATERIALS PROVIDED

Description	Code	Quantity
NRG 1-β1 Microplate - 96 well polystyrene microplate (12 strips of 8 wells) coated with a mouse monoclonal antibody against NRG 1-β1.	556-01-01	1 plate
NRG 1-β1 Standard – 4000 pg/vial of recombinant human NRG 1-β1 in a buffered protein base with preservatives; lyophilized.	556-01-02	2 vials
Detection Antibody Concentrate – 105 µL/vial, 100-fold concentrate of biotinylated polyclonal antibody against NRG 1-β1 with preservatives; lyophilized.	556-01-03	1 vial
Positive Control – one vial of recombinant human NRG 1-β1, lyophilized	556-01-04	1 vial
Streptavidin-HRP Conjugate - 120 µL/vial, 100-fold concentrated solution of Streptavidin conjugate to HRP with preservatives	SAHRP	1 vial
Dilution Buffer - 60mL of buffered protein based solution with preservatives	DB08	1 bottle
Wash Buffer - 50 mL of 10-fold concentrated buffered surfactant, with preservative.	WB01	1 bottle
TMB Substrate Solution -11 mL of TMB substrate solution	TMB01	1 bottle
Stop Solution - 11 mL of 0.5M HCl	S-STOP	1 bottle
Plate Sealer	EAPS	1
Plastic Pouch	P01	1

STORAGE

Unopened Kit: Store at 2 - 8°C for up to 12 months. For longer storage, unopened Standard, Positive Control and Detection Antibody Concentrate should be stored at -20°C or -70°C. Do not use kit past expiration date.

Opened / Reconstituted Reagents: Reconstituted Standard and Detection Antibody Concentrate Solution SHOULD BE STORED at -20°C or -70°C for up to one month. Streptavidin-HRP Conjugate 100-fold concentrate and other components may be stored at 2 - 8°C for up to 12 months.

Microplate Wells: Return unused wells to the plastic pouch with the desiccant pack and seal along entire edge of zip-seal. Microplate may be stored for up to 6 months at 2 - 8°C after opening.

OTHER SUPPLIES REQUIRED

- Microplate reader capable of measuring absorbance at 450 nm, with the correction wavelength set at 540 nm or 570 nm.
- Microplate shaker (250-300rpm).
- Pipettes and pipette tips.
- Deionized or distilled water.
- Squirt bottle, manifold dispenser, or automated microplate washer.
- 100 mL and 500 mL graduated cylinders.

SAMPLE COLLECTION AND STORAGE

Serum - Use a serum separator tube (SST) and allow samples to clot for 30 minutes before centrifugation for 15 minutes at 1000 x g. Remove serum and assay immediately or aliquot and store samples at ≤ -20° C. Avoid repeated freeze-thaw cycles.

Plasma - Collect plasma using EDTA, heparin, or citrate as an anticoagulant. Centrifuge for 15 minutes at 1000 x g within 30 minutes of collection. Assay immediately or aliquot and store samples at ≤ -20° C. Avoid repeated freeze-thaw cycles.

Note: Use Aprotinin (enzyme inhibitor) (Code No.: 00700-01-25) for ALL sample collection to prevent sample degradation. 0.5 TIU per ml of sample solution.

SAMPLE PREPARATION

*Because sNRG1-beta1 levels in serum or plasma samples vary widely, **optimal dilutions should be determined by each laboratory for each application with a pretest.***

It is recommended to directly assay all samples first. If samples levels are less than 250 pg/mL lyophilize 500 µL of samples and reconstitute it with 250 µL of Dilution Buffer, this will concentrate that sample by 2 times. If sample levels are higher than 8000 pg/mL, it may require an 8-fold or higher dilution. A suggested 8-fold dilution is 30 µL sample + 210 µL Dilution Buffer.

Use polypropylene test tubes.

This ELISA kit has two vials of standard for repeat assay with optimal dilution factors.

PRECAUTIONS FOR USE

All reagents should be considered as potentially hazardous. The stop solution contains diluted hydrochloric acid. Appropriate care, therefore, should be taken while handling this solution. We therefore recommend that this product is handled only by those persons who have been trained in laboratory techniques and that it is used in accordance with the principles of good laboratory practice. Wear suitable protective clothing such as laboratory overalls, safety glasses and gloves. Care should be taken to avoid contact with skin or eyes. In the case of contact with skin or eyes wash immediately with water.


REAGENT PREPARATION

Bring all reagents to room temperature before use.

Wash Buffer - If crystals have formed in the concentrate, warm to room temperature and mix gently until the crystals have completely dissolved. Dilute 50 mL of Wash Buffer Concentrate into deionized or distilled water (450 mL) to prepare 500 mL of Wash Buffer.

NRG 1-β1 Standard - Refer to vial label for reconstitution volume. Reconstitute the **NRG 1-β1** standard with **0.5 mL** of Dilution Buffer. This reconstitution produces a stock solution of 8000 pg/mL. Allow the standard to sit for a minimum of 15 minutes with gentle agitation prior to making dilutions. Pipette 250µL of the appropriate Dilution Buffer into tubes #1 to #6. Use the stock solution to produce a dilution series (below). Mix each tube thoroughly before the next transfer. The 8000 pg/mL standard serves as the high standard. The appropriate Dilution Buffer serves as the zero standard (0 pg/mL).

Tube	Standard	Dilution Buffer	Concentration
Stock	Powder	500 µl	8000 pg/ml
# 1	250 µl of stock	250 µl	4000 pg/ml
# 2	250 µl of 1	250 µl	2000 pg/ml
# 3	250 µl of 2	250 µl	1000 pg/ml
# 4	250 µl of 3	250 µl	500 pg/ml
# 5	250 µl of 4	250 µl	250 pg/ml
# 6	250 µl of 5	250 µl	125 pg/ml



Positive Control - Reconstitute the positive control with 500 µL of Dilution Buffer to make positive control solution. **Note:** Positive Control should be used immediately.

Detection Antibody Concentrate - Reconstitute the Detection Antibody Concentrate with 105 µL of Dilution Buffer to produce a 100-fold concentrated stock solution. Transfer 105 µL of 100-fold concentrated stock solution to 10.395 mL of Dilution Buffer to prepare working solution.

Streptavidin-HRP Conjugate - Pipette 11.88 mL of Dilution Buffer into a 15 mL centrifuge tube and transfer 120 µL of 100-fold concentrated stock solution to prepare working solution. **Note:** 1x working solution of Streptavidin-HRP Conjugate should be used within a few days.


ASSAY PROCEDURE

Bring all reagents and samples to room temperature before use. It is recommended that blank, standards, positive control and samples be assayed in duplicates.

1. Prepare all reagents and working standards as directed in the previous sections.
2. Remove excess micro-plate strips from the plate frame, return them to the plastic pouch with the desiccant pack and seal.
3. Add 100 µL of Dilution Buffer to Blank wells (F4, F5).
4. Add 100 µL of Standard (from B2, B3 to G2, G3 and G4, G5), sample, or positive control (E4, E5) per well. Cover with plate sealer. Incubate for 2 hours on micro-plate shaker at room temperature. A plate layout is provided to record standards and samples assayed.
5. Aspirate each well and wash, repeating the process three times for a total of four washes.

Wash by filling each well with 1x Wash Buffer (300 µL) using a squirt bottle, manifold dispenser, or autowasher. Complete removal of liquid at each step is essential to good performance. After the last wash, remove any remaining Wash Buffer by aspirating or decanting. Invert the plate and blot it against clean paper towels.

6. Add 100 µL of Detection Antibody working solution to each well. Cover with plate sealer. Incubate for 2 hours on micro-plate shaker at room temperature.
7. Repeat the aspiration/wash as in step 5.
8. Add 100 µL of Streptavidin-HRP Conjugate working solution to each well. Incubate for 1 hour on micro-plate shaker at room temperature. **Protect from light.**
9. Repeat the aspiration/wash as in step 5.
10. Add 100 µL of Substrate Solution to each well. Incubate for 25-35 minutes at room temperature. **Protect from light.**
11. Add 100 µL of Stop Solution to each well. The color in the wells should change from blue to yellow. If the color in the wells is green, or if the color change does not appear uniform, gently tap the plate to ensure thorough mixing. It is recommended to add the stop solution when the highest standard has developed a dark blue color.
12. Determine the optical density of each well within 15 minutes, using a micro-plate reader set to 450 nm.



CALCULATION OF RESULTS

Average the duplicate readings for each standard, positive control, and sample and subtract the average zero standard optical density. Create a standard curve by reducing the data using computer software capable of generating a log-log curve fit. As

an alternative, construct a standard curve by plotting the mean absorbance for each standard on the y-axis against the concentration on the x-axis and draw a best fit curve through the points on the graph. The data may be linearized by plotting the log of the NRG 1-β1 concentrations versus the log of the O.D. and the best fit line can be determined by regression analysis. This procedure will produce an adequate but less precise fit of the data.

If samples have been diluted, the concentration read from the standard curve must be multiplied by the dilution factor.

TYPICAL DATA

This standard curve is provided for demonstration only. A standard curve should be generated for each set of samples assayed.

STANDARD (pg/mL)	CORRECTED OD450nm
Blank	0 (0.087)
125	0.015
250	0.029
500	0.055
1000	0.118
2000	0.258
4000	0.540
8000	0.885

- Lot No.:
- Positive Control: 1000 – 2500 pg/mL

CALIBRATION

This immunoassay is calibrated against a highly purified *E. Coli*-expressed recombinant human NRG 1-β1, ECD.

SENSITIVITY

62.5 pg/mL.

SPECIFICITY

This assay recognizes both natural and recombinant human NRG 1-β1. The factors listed below were prepared at 50 ng/mL in Dilution Buffer, and assayed for cross reactivity. Preparations of the following factors at 50 ng/mL in a mid-range rh NRG 1-β1 control were assayed for interference. No significant cross-reactivity or interference was observed.

PROTEINS	CROSS-REACTIVITY (%)
Human NRG 1-β1	100
Human ErbB2/Fc chimera	0
Human ErbB3/Fc chimera	0
Human NRG4 extracellular domain	0
Human NRG 1 –α (EGF domain)	0
Human NRG1 isoform SMDF	0

LINEARITY









To assess the linearity of the assay, pooled research human **serum** samples were diluted with Dilution Buffer and assayed.

DILUTION FACTOR	ASSAYED (pg/mL)	FINAL (pg/mL)	RECOVERY (%)
4x	7318.906	29275.624	100
8x	3467.657	27741.256	94.8
16x	1645.224	26323.584	89.9

To assess the linearity of the assay, pooled research human **EDTA plasma** samples were diluted with Dilution Buffer and assayed.

DILUTION FACTOR	ASSAYED (pg/mL)	FINAL (pg/mL)	RECOVERY (%)
4x	6001.472	24005.888	100
8x	3078.534	24628.272	103
16x	1404.607	22473.712	93.6

SUMMARY OF ASSAY PROCEDURE**PREPARE REAGENTS, SAMPLES AND STANDARDS**


Add 100 µl of standard, samples, positive control to each well. Incubate 2 hours on the plate shaker at RT.

Aspirate and wash 4 times.

Add 100 µl Detection Antibody working solution to each well. Incubate 2 hours on the plate shaker at RT.

Aspirate and wash 4 times.

Add 100 µl Streptavidin-HRP conjugate working solution to each well. Incubate 1 hour on the plate shaker at RT. Protect from light.

Aspirate and wash 4 times.

Add 100 µl Substrate Solution to each well. Incubate 25-35 min on the plate shaker. Protect from light.

Add 100 µl Stop Solution to each well. Read 450nm within 15 min