

HUMAN SOLUBLE GLYCOPROTEIN OF 130 KDA (sgp 130) ELISA KIT

FOR THE QUANTITATIVE DETERMINATION OF HUMAN SGP 130 CONCENTRATIONS IN CELL CULTURE SUPERNATES, SERUM, AND PLASMA.



FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.

PURCHASE INFORMATION:

ELISA Name	HUMAN sgp130 ELISA
Catalog No.	SK00155-01
Lot No.	
Formulation	96 T
Standard range	78-10000 pg/mL
Sensitivity	39 pg/mL
Sample Volume	100 µl
Dilution factor	50 (<i>Optimal dilutions should be determined by each laboratory for each application</i>)
Sample Type	Serum, EDTA Plasma, Cell Culture Supernates
Specificity	Human sgp 130
Intra-assay Precision	4-6%
Inter-assay Precision	8-10%
Storage	2-8 °C

Order Contact:
 AVISCERA BIOSCIENCE INC.
 2348 Walsh Ave., Suite C
 Santa Clara, CA 95051
 Tel: (408) 982 0300
 Fax: (408) 982 0301
 Email: Sales@AvisceraBioscience.com
 Info@AvisceraBioscience.com
www.AvisceraBioscience.com

INTRODUCTION

Human sgp 130 immunoassay is a 3.5 - 4.5 hour solid phase ELISA designed to measure human sgp 130 in cell culture supernates, serum, and plasma. It contains recombinant human sgp 130 and antibodies raised against this protein. It has been shown to accurately quantify recombinant human sgp 130. Results obtained with naturally occurring sgp 130 samples showed linear curves that were parallel to the standard curves obtained using the kit standards. These results indicate that the immunoassay kit can be used to determine relative mass values for natural human sgp 130.

PRINCIPLE OF THE ASSAY

This assay employs the quantitative sandwich enzyme immunoassay technique. A monoclonal antibody specific for sgp 130 has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and any sgp 130 present is bound by the immobilized antibody. After washing away any unbound substances, a biotinylated polyclonal antibody specific for sgp 130 is added to the wells. Following a wash to remove any unbound antibody-biotin reagent, HRP link Streptavidin is added to the wells. After washing away any unbound enzyme, a substrate solution is added to the wells and color develops in proportion to the amount of sgp 130 bound in the initial step. The color development is stopped and the intensity of the color is measured.

LIMITATIONS OF THE PROCEDURE

_ FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.

_ The kit should not be used beyond the expiration date on the kit label.

_ Do not mix or substitute reagents with those from other lots or sources.

_ It is important that the Dilution Buffer selected for the standard curve be consistent with the samples being assayed.

_ If samples generate values higher than the highest standard, dilute the samples with Dilution Buffer and repeat the assay.

_ Any variation in standard diluent, operator, pipetting technique, washing technique, incubation time or temperature, and kit age can cause variation in binding.

_ This assay is designed to eliminate interference by soluble receptors, binding proteins, and other factors present in biological samples. Until all factors

have been tested in the immunoassay, the possibility of interference cannot be excluded.

MATERIALS PROVIDED

Description	Code	Quantity
sgp 130 Microplate - 96 well polystyrene microplate (12 strips of 8 wells) coated with a monoclonal antibody against sgp 130.	155-01-01	1 plate
sgp 130 Standard – 10000pg/vial of recombinant human sgp 130 in a buffered protein base with preservatives; lyophilized.	155-01-02	1 vial
Detection Antibody Concentrate – 105 µL/vial, 100-fold concentrated of biotinylated polyclonal antibody against sgp 130 with preservatives; lyophilized.	155-01-03	1 vial
Positive Control - one vial of recombinant human sgp 130, lyophilized	155-01-04	1 vial
Streptavidin-HRP Conjugate - 60 uL/vial, 200-fold concentrated solution of Streptavidin conjugate to HRP with preservatives	SAHRP	1 vial
Dilution Buffer - 60mL of buffered protein based solution with preservatives	DB01	2 bottles
Antibody Diluent Solution Concentrate – 11mL of buffered protein based solution with preservatives	DB20	1 tube
Wash Buffer - 50 mL of 10-fold concentrated buffered surfactant, with preservative.	WB01	1 bottle
TMB Substrate Solution - 11 mL of TMB substrate solution	TMB01	1 bottle
Stop Solution – 11mL of 0.5M HCl	S-STOP	1 bottle
Plate Sealer	EAPS	1 piece
Plastic Pouch	P01	1 piece

STORAGE

Unopened Kit: Store at 2 - 8°C for up to 12 months. For longer storage, unopened Standard, Positive Control, Detection Antibody Concentrate and Antibody Diluent Solution Concentrate should be

stored at -20°C or -70°C. Do not use kit past expiration date.

Opened / Reconstituted Reagents: Reconstituted Standard, Detection Antibody Concentrate Solution and Antibody Diluent Solution SHOULD BE STORED at -20°C or -70°C for up to one month. Streptavidin-HRP Conjugate 200-fold concentrate and other components may be stored at 2 - 8°C for up to 12 months.

Microplate Wells: Return unused wells to the plastic pouch with the desiccant pack and seal along entire edge of zip-seal. Microplate may be stored for up to 6 months at 2 - 8°C after opening.

OTHER SUPPLIES REQUIRED

- Microplate reader capable of measuring absorbance at 450 nm, with the correction wavelength set at 540 nm or 570 nm.
- Microplate shaker (250-300rpm).
- Pipettes and pipette tips.
- Deionized or distilled water.
- Squirt bottle, manifold dispenser, or automated microplate washer.
- 100 mL and 500 mL graduated cylinders.

SAMPLE COLLECTION AND STORAGE

Cell Culture Supernates - Remove particulates by centrifugation and assay immediately or aliquot and store samples at ≤-20° C. Avoid repeated freeze-thaw cycles.

Serum - Use a serum separator tube (SST) and allow samples to clot for 30 minutes before centrifugation for 15 minutes at 1000 x g. Remove serum and assay immediately or aliquot and store samples at ≤ -20° C. Avoid repeated freeze-thaw cycles.

Plasma - Collect plasma using EDTA, heparin, or citrate as an anticoagulant. Centrifuge for 15 minutes at 1000 x g within 30 minutes of collection. Assay immediately or aliquot and store samples at ≤-20° C. Avoid repeated freeze-thaw cycles.

Note: Use Aprotinin (enzyme inhibitor) (Code No.: 00700-01-25) for ALL sample collection to prevent sample degradation. 0.5 TIU per ml of sample solution.

SAMPLE PREPARATION

Serum and plasma samples may require a 50-fold dilution. A suggested 50-fold dilution is 5µL sample + 245 µL Dilution Buffer. **Optimal dilutions should be determined by each laboratory for each application. Use polypropylene test tubes.**


REAGENT PREPARATION

Bring all reagents to room temperature before use.

Wash Buffer - If crystals have formed in the concentrate, warm to room temperature and mix gently until the crystals have completely dissolved. Dilute 50 mL of Wash Buffer Concentrate into deionized or distilled water (450 mL) to prepare 500 mL of Wash Buffer.

sgp 130 Standard - Refer to vial label for reconstitution volume. Reconstitute the **sgp 130** standard with 1.0 mL of Dilution Buffer. This reconstitution produces a stock solution of 10000 pg/mL. Allow the standard to sit for a minimum of 15 minutes with gentle agitation prior to making dilutions. Pipette 250 µL of Dilution Buffer into tubes #1 to #7. Use the stock solution to produce a dilution series (below). Mix each tube thoroughly before the next transfer. The 10000 pg/mL standard serves as the high standard. The Dilution Buffer serves as the zero standard (0 pg/mL).

Tube	Standard	Dilution Buffer	Concentration
stock	Powder	1.0 ml	10000 pg/ml
# 1	250µl of stock	250µl	5000 pg/ml
# 2	250µl of 1	250µl	2500 pg/ml
# 3	250µl of 2	250µl	1250 pg/ml
# 4	250µl of 3	250µl	625 pg/ml
# 5	250µl of 4	250µl	312.5 pg/ml
# 6	250µl of 5	250µl	156 pg/ml
# 7	250µl of 6	250µl	78 pg/ml



Antibody Diluent Solution Concentrate –

Reconstitute the Antibody Diluent Solution Concentrate with 11.0 mL of Dilution Buffer in provided 15 mL tube to prepare Antibody Diluent Solution.

Detection Antibody - Reconstitute the Detection Antibody Concentrate with 105 µl of **Antibody Diluent Solution (DB20)** to produce a 100-fold concentrated stock solution. Pipette 10.395 mL of Antibody Diluent Solution into another 15 ml centrifuge tube and transfer the 105 µl of 100-fold concentrated stock solution to prepare working solution. **Note: Must be prepared 1 to 2 hours prior to use.**

Streptavidin-HRP Conjugate - Pipette 11.94 mL of Dilution Buffer into a 15 ml centrifuge tube and transfer 60 µl of 200-fold concentrated stock solution to prepare working solution. **Note:** 1x working solution of Streptavidin-HRP Conjugate should be used within a few days.


Positive Control - Reconstitute the positive control with 1.0 mL of Dilution Buffer to make positive control solution. **Note:** Positive Control should be used immediately.

ASSAY PROCEDURE

Bring all reagents and samples to room temperature before use. It is recommended that blank, positive control, standards and samples be assayed in duplicates.

1. Prepare all reagents and working standards as directed in the previous sections.
2. Remove excess micro-plate strips from the plate frame, return them to the plastic pouch with the desiccant pack.
3. Add 100 µL of **Dilution Buffer** to Blank wells (B2, B3).
4. Add 100 µL of **Standard** (D2, D3 to G2, G3 and D4, D5 to G4, G5), **sample**, or **positive control** (C2, C3) per well. Cover with plate sealer. Incubate for 2 hours on micro-plate shaker at room temperature. A plate layout is provided to record standards and samples assayed.
5. Aspirate each well and wash, repeating the process three times for a total of four washes. Wash by filling each well with **Wash Buffer** (300 µL) using a squirt bottle, manifold dispenser, or autowasher. Complete removal of liquid at each step is essential to good performance. After the last wash, remove any remaining Wash Buffer by aspirating or decanting. Invert the plate and blot it against clean paper towels.

6. Add 100 µL of **Detection Antibody working solution** to each well. Cover with plate sealer. Incubate for 2 hours on micro-plate shaker at room temperature.
7. Repeat the aspiration/wash as in step 5.
8. Add 100 µL of **Streptavidin-HRP Conjugate working solution** to each well. Incubate for 60 minutes on micro-plate shaker at room temperature. **Protect from light.**
9. Repeat the aspiration/wash as in step 5.
10. Add 100 µL of **Substrate Solution** to each well. Incubate for 2-4 minutes at room temperature. **Protect from light.**
11. Add 100 µL of **Stop Solution** to each well. The color in the wells should change from blue to yellow. If the color in the wells is green, or if the color change does not appear uniform, gently tap the plate to ensure thorough mixing.
12. Determine the optical density of each well within 15 minutes, using a micro-plate reader set to 450 nm.



CALCULATION OF RESULTS

Average the duplicate readings for each standard, positive control, and sample and subtract the average zero standard optical density. Create a standard curve by reducing the data using computer software capable of generating a log-log curve fit. As an alternative, construct a standard curve by plotting the mean absorbance for each standard on the y-axis against the concentration on the x-axis and draw a best fit curve through the points on the graph. The data may be linearized by plotting the log of the sgp130 concentrations versus the log of the O.D. and the best fit line can be determined by regression analysis. This procedure will produce an adequate but less precise fit of the data.

If samples have been diluted, the concentration read from the standard curve must be multiplied by the dilution factor.

CALIBRATION

This immunoassay is calibrated against a highly purified Sf21-expressed recombinant human sgp 130.

SENSITIVITY

Twenty-five assays were evaluated and the minimum detectable dose (MDD) of SGP 130 was 39 pg/mL.

SPECIFICITY

The following factors prepared at 50 ng/mL were assayed and exhibited no cross-reactivity or interference.

Protein	Cross-reactivity (%)
Human sgp 130	100
Mouse sgp 130	0
Human CNTF	0
Human IL-6	0
Human IL-11	0

LINEARITY

To assess the linearity of the assay, pooled research human EDTA plasma samples were diluted with Dilution Buffer and assayed.

DILUTION FACTOR	ASSAYED (PG/ML)	FINAL (PG/ML)	RECOVERY (%)
50X	3847	192350	100
100X	2053	205300	107

To assess the linearity of the assay, pooled research human serum samples were diluted with Dilution Buffer and assayed.

DILUTION FACTOR	ASSAYED (PG/ML)	FINAL (PG/ML)	RECOVERY (%)
50X	4079	203950	100
100X	2396	239600	117

TYPICAL DATA

This standard curve is provided for demonstration only. A standard curve should be generated for each set of samples assayed.

STANDARD (PG/ML)	AVERAGE OD450 (CORRECTED)*
Blank	0 (0.068)
78	0.028
156	0.058
312.5	0.127
625	0.235
1250	0.450
2500	0.785
5000	1.395
10000	2.134

*Lot No.: 20110790

** Positive Control: 700 – 1500 pg/mL

SUMMARY OF ASSAY PROCEDURE

