

# COMPLEMENT COMPONENT C2 (HUMAN) ELISA KIT

FOR THE QUANTITATIVE DETERMINATION OF HUMAN COMPLEMENT COMPONENT C2 CONCENTRATIONS IN SERUM



ALWAYS REFER TO LOT SPECIFIC PROTOCOL PROVIDED WITH EACH KIT FOR INSTRUCTIONS. PROTOCOL MUST BE READ BEFORE USING THIS PRODUCT.

FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.

## PRODUCT INFORMATION:

ELISA NAME	COMPLEMENT COMPONENT C2 (HUMAN) ELISA KIT
Catalog No.	SK00165-01
Formulation	96 T
Lot No.	
Standard range	234.4-15000 pg/mL
Sensitivity	160 pg/mL
Sample Volume	100 µL
Dilution Factor	Optimal dilutions should be determined by each laboratory for each application
Sample Type	Serum
Specificity	Human C2
Calibration	Human C2 rec
Intra-assay Precision	4 - 6%
Inter-assay Precision	8 - 10%
Storage	2 – 8 °C
This kit contains sufficient materials to run 35 samples duplicated provided that assay is run according to protocol.	

Order Contact:  
**AVISCIERA BIOSCIENCE, INC.**  
 2348 Walsh Ave., Suite C  
 Santa Clara, CA 95051  
 Tel: (408) 982 0300  
 Fax: (408) 982 0301  
 Email: [Sales@AvisceraBioscience.com](mailto:Sales@AvisceraBioscience.com)  
[Info@AvisceraBioscience.com](mailto:Info@AvisceraBioscience.com)  
[www.AvisceraBioscience.com](http://www.AvisceraBioscience.com)

**DESCRIPTION**

This Complement Component C2 (Human) ELISA Kit contains the necessary components required for the quantitative measurement of recombinant and/or natural human Complement Component C2 from serum in a sandwich ELISA format.

This immunoassay contains recombinant Complement Component C2 (Human) and antibodies raised against this protein. Results from this immunoassay have shown to accurately quantify recombinant and natural the Complement Component C2 samples.

**ASSAY OVERVIEW**

This assay employs the quantitative sandwich ELISA format. The plate is pre-coated with a monoclonal antibody specific for Complement Component C2 (Human). The capture antibody can bind to the mouse gp130 in the standard and samples. After washing the plate of any unbound substances, a biotinylated monoclonal antibody against Complement Component C2 (Human) is added to the wells. After another washing of the plate, the streptavidin-HRP Conjugate is added. After the last wash to remove any unbound enzyme, a substrate solution (TMB) is added to the wells and color develops in direct proportion to the amount of Complement Component C2 (Human) bound in the standard solutions or samples. A standard curve can be established and sample values can be read off the standard curve.

**COMPONENTS PROVIDED**

DESCRIPTION	CODE	QUANTITY
<b>C2 Microplate</b> – 96 well microplate coated with antibody specific for human Complement Component C2.	<b>165-01-01</b>	<b>1 plate</b>
<b>C2 Standard</b> – 15000 pg/vial of lyophilized recombinant human C2.	<b>165-01-02</b>	<b>1 vial</b>
<b>Detection Antibody HRP Concentrate</b> – 0.105mL/vial of 100-fold concentrate of lyophilized biotinylated antibody against human C2.	<b>165-01-03</b>	<b>1 vial</b>

<b>Positive Control</b> – one vial of lyophilized recombinant human C2.	<b>165-01-04</b>	<b>1 vial</b>
<b>Dilution Buffer</b> – 60 mL of buffered solution with preservative.	<b>DB10</b>	<b>1 bottle</b>
<b>Wash Buffer</b> – 50 mL of 10-fold concentrated buffered surfactant with preservative.	<b>WB01</b>	<b>1 bottle</b>
<b>TMB Substrate Solution</b> – 11 mL of TMB substrate solution.	<b>TMB01</b>	<b>1 bottle</b>
<b>Stop Solution</b> – 11 mL of 0.5M HCl.	<b>S-STOP</b>	<b>1 bottle</b>
<b>Plate Sealer</b>	<b>EAPS</b>	<b>1 piece</b>
<b>Plastic Pouch</b>	<b>P01</b>	<b>1 piece</b>

**PROCEDURAL LIMITATIONS**

\_FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.

\_This ELISA kit should not be used beyond the expiration date on the kit label.

\_Do not mix reagents with those from other lots or sources.

\_It is important that the Dilution Buffer selected for the standard curve be consistent with the samples being assayed.

\_Each laboratory must determine the optimal dilution factors for the samples being assayed with a pretest. If samples generate values that are not within the dynamic range of the standard curve, further concentrate or dilute the samples as required with Dilution Buffer and repeat the assay  
\_Any modifications in buffers, pipetting technique, washing technique, incubation time or temperature, as well as kit age can cause a change in signal.

\_Not all interfering factors have been tested in the immunoassay, therefore the possibility of interference cannot be excluded.

**STORAGE**

**Unopened Kit:** Store at 2 – 8 °C for up to 8 months. For longer storage, unopened Standard, Positive

Control and Detection Antibody Concentrate should be stored at -20 °C or -70 °C. Do not use kit past expiration date.

**Opened / Reconstituted Reagents:** Reconstituted Detection Antibody concentrated solution SHOULD BE STORED at -20 °C or -70 °C for up to one month. Streptavidin-HRP Conjugate 100-fold concentrated solution and TMB Substrate Solution can be stored at 2 – 8 °C for up to 8 months (**DO NOT FREEZE and PROTECT FROM LIGHT**). All other components may be stored at 2 – 8 °C for up to 8 months.

**Microplate Wells:** Return unused strips to the plastic pouch with the desiccant pack. Microplate may be stored for up to 6 months at 2 – 8 °C after opening.

**ADDITIONAL MATERIALS REQUIRED**

- Microplate reader capable of absorbance measurement at 450 nm.
- Microplate shaker (250 – 300 rpm).
- Microplate washer or manifold dispenser.
- 100 mL and 500 mL graduated cylinders.
- Multi-channel Pipette, Pipettes and pipette tips.
- Deionized or distilled water.

**PRECAUTION**

This kit should be handled by those persons who have been trained in and can follow the principles of good laboratory practice. Wear protective clothing such as laboratory overalls, safety glasses and gloves. Care should be taken while handling solutions in this kit to avoid contact with skin or eyes, especially with the stop solution because it contains diluted hydrochloric acid. Wash immediately with water in case of contact on skin or eyes.

**SAMPLE COLLECTION AND STORAGE**

**Serum** - Use a serum separator tube (SST) and allow samples to clot for 30 minutes before centrifugation for 15 minutes at 1000 x g. Remove serum and assay immediately or aliquot and store samples at ≤ -20° C. Avoid repeated freeze-thaw cycles.

Optional: Use Protease Inhibitors (Order No.: PI-00978-50) for ALL sample collection to prevent sample degradation. 10 µL per ml of sample solution.

**SAMPLE PREPARATION**

**Optimal dilutions should be determined by each laboratory for each application with a pretest.**

**Use polypropylene test tubes.**

**REAGENT PREPARATION**

**Bring all reagents to room temperature before use.**

**Wash Buffer** – Dilute 50 mL of Wash Buffer Concentrate into 450 mL distilled or deionized water to make 500 mL of 1x Wash Buffer. If crystals have formed in the concentrate, warm bottle in a water bath until the crystals have completely dissolved.

**Human C2 Standard** – Reconstitute the human C2 standard with 1.0 mL of Dilution Buffer. The concentration of the reconstituted stock solution is 15000 pg/mL. Allow the stock standard to sit for at least 15 minutes with gentle agitation until completely dissolved prior to making standard dilutions (see below). Mix each tube thoroughly before the next transfer.

TUBE	STANDARD	DILUTION BUFFER	CONCENTRATION
Stock	powder	1.0 mL	15000pg/mL
# 1	250µL of stock	250µL	7500 pg/mL
# 2	250µL of 1	250µL	3750 pg/mL
# 3	250µL of 2	250µL	1875 pg/mL
# 4	250µL of 3	250µL	937.5 pg/mL
# 5	250µL of 4	250µL	468.75 pg/mL
# 6	250µL of 5	250µL	234.4 pg/mL

**Detection Antibody** - Pipette 10.395 mL of Dilution Buffer into a 15 mL centrifuge tube and transfer 0.105 mL of 10-fold concentrated stock solution to prepare working solution. (**protect from light**). **DO NOT FREEZE.**

**Positive Control** - Reconstitute the Positive Control with 1.0 mL Dilution Buffer. **Note:** Positive Control could be used within a few days if stored at -20 °C or -70 °C.

**ELISA PROTOCOL**

**Bring all reagents and samples to room temperature before the start of the assay. Blank, standard dilutions, positive control and samples should be assayed in duplicate. ELISA Protocol may need further optimization.**

1. Prepare all reagents, standard dilutions, positive control and samples as directed previously.
2. Remove unneeded microplate strips from the plate frame and return them to the plastic pouch with the desiccant pack.
3. Add 100 µL per well of **Dilution Buffer** to Blank wells.
4. Add 100 µL per well of **Standard Dilutions, sample, or positive control**. Cover with plate sealer and incubate for 2 hours on microplate shaker at room temperature. (Please see plate layout provided.)
5. Aspirate and wash each well with 300 µL of **1x Wash Buffer** four times. After the last wash, aspirate any remaining 1x Wash Buffer, invert the plate and blot against clean paper towel(s).
6. Add 100 µL per well of **Detection Antibody working solution**. Cover with plate sealer and incubate for 2 hours on microplate shaker at room temperature.
7. Repeat the aspiration and wash as in step 5.
8. Add 100 µL per well of **Streptavidin-HRP Conjugate working solution**. Cover with plate sealer and incubate for 50 minutes on microplate shaker at room temperature. **Protect from light.**
9. Repeat the aspiration and wash as in step 5.
10. Add 100 µL per well of **Substrate Solution**. Incubate for 5-10 minutes on microplate shaker at room temperature. **Protect from light.**
11. Add 100 µL per well of **Stop Solution**. The color in the wells should change from blue to yellow. If the color change does not appear uniform, gently tap the plate to ensure thorough mixing.
12. Read plate using a microplate reader set to 450 nm within 15 minutes.

### SPECIFICITY

PROTEIN	CROSS-REACTIVITY
Human C2	100%
Human C5a	0
Human Adipsin	0

### TYPICAL STANDARD CURVE

This standard curve is for demonstration only. A new standard curve should be made for each set of samples assayed.







STANDARD (PG/ML)	AVERAGE OD450 (CORRECTED)
Blank	0 (0.096)
234.4	0.041
468.75	0.083
937.5	0.157
1875	0.319
3750	0.621
7500	1.220
15000	2.440

### CALCULATION OF RESULTS

Create a standard curve by plotting the log of the known concentrations of the standard dilutions (x-axis) versus the log of its corresponding O.D. (y-axis) and draw the best fit line through the points. It is recommended to use computer software capable of generating a log-log curve fit to more accurately quantify the standard dilutions.

If samples have been diluted, the concentration read from the standard curve must be multiplied by the dilution factor.

**SUMMARY OF ASSAY PROCEDURE**

<b>PREPARE REAGENTS, SAMPLES AND STANDARD DILUTIONS</b>

Add 100 µL of standard dilutions, samples and positive control. Cover with plate sealer and incubate 2 hours on microplate shaker at RT.

Aspirate and wash 4 times.

Add 100 µL per well of Detection Antibody working solution. Cover with plate sealer and incubate 1 hour on microplate shaker at RT.

Aspirate and wash 4 times.

Add 100 µL per well of Substrate Solution. Incubate 5-10 min on microplate shaker at RT. <b>Protect from light.</b>

Add 100 µL per well of Stop Solution. Read at 450 nm within 15 minutes.