

HUMAN IL-17A ELISA KIT

FOR THE QUANTITATIVE DETERMINATION OF HUMAN IL-17A CONCENTRATIONS IN CELL CULTURE SUPERNATES, SERUM AND PLASMA



ALWAYS REFER TO LOT SPECIFIC PROTOCOL PROVIDED WITH EACH KIT FOR INSTRUCTIONS. PROTOCOL MUST BE READ BEFORE USING THIS PRODUCT.

FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.

PRODUCT INFORMATION:

THIS IS FOR ONE TIME USE.

ELISA NAME	HUMAN IL-17A ELISA
Catalog No.	SK00730-01
Lot No.	
Formulation	96 T
Standard Range	3.9 - 250 pg/ml
Sensitivity	3 pg/ml
Sample Volume	100 µl
Dilution Factor	Optimal dilutions should be determined by each laboratory for each application
Sample Type	Serum, EDTA Plasma, Cell Culture Supernates
Specificity	Human IL-17A
Calibration	Human IL-17A recombinant
Intra-assay Precision	4 - 10%
Inter-assay Precision	8 - 12%
Storage	2 – 8° C
This kit contains sufficient materials to run 35 samples duplicated provided that assay is run according to protocol.	

Order Contact:

AVISCIERA BIOSCIENCE, INC.

2348 Walsh Ave., Suite C

Santa Clara, CA 95051

USA

Tel: (408) 982 0300

Fax: (408) 982 0301

Email: Sales@AvisceraBioscience.com

Info@AvisceraBioscience.com

www.AvisceraBioscience.com

DESCRIPTION

This Human IL-17A ELISA Kit contains the necessary components required for the quantitative measurement of recombinant and/or natural human IL-17A from cell culture supernates, serum and plasma in a sandwich ELISA format.

This immunoassay contains recombinant human IL-17A and antibodies raised against this protein. Results from this immunoassay have shown to accurately quantify recombinant and natural IL-17A samples.

ASSAY OVERVIEW

This assay employs the quantitative sandwich ELISA format. The plate is pre-coated with an antibody specific for human IL-17A. The capture antibody can bind to the human IL-17A in the standard and samples. After washing the plate of any unbound substances, a biotinylated antibody against human IL-17A is added to the wells. After another washing of the plate, Avidin-HRP Conjugate is added. After the last wash to remove any unbound enzyme, a substrate solution is added to the wells and color develops in direct proportion to the amount of human IL-17A bound in the standard solutions or samples. A standard curve can be established and sample values can be read off the standard curve.

PROCEDURAL LIMITATIONS

_FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.

_This ELISA kit should not be used beyond the expiration date on the kit label.

_Do not mix reagents with those from other lots or sources.

_It is important that the Dilution Buffer selected for the standard curve be consistent with the samples being assayed.

_Each laboratory must determine the optimal dilution factors for the samples being assayed with a pretest. If samples generate values that are not within the dynamic range of the standard curve, further concentrate or dilute the samples as required with Dilution Buffer and repeat the assay.

_Any modifications in buffers, pipetting technique, washing technique, incubation time or temperature, as well as kit age can cause a change in signal.

_Not all interfering factors have been tested in the immunoassay, therefore the possibility of interference cannot be excluded.

COMPONENTS PROVIDED

DESCRIPTION	CODE	QUANTITY
IL-17A Microplate - 96 well polystyrene microplate coated with an antibody against IL-17A.	730-01-01	1 plate
IL-17A Standard – lot specific of recombinant human IL-17A in a buffered protein base with preservative; lyophilized.	730-01-02	1 vial
Detection Antibody Concentrate – lot specific concentrate of biotinylated antibody against IL-17A with preservative; lyophilized.	730-01-03	1 vial
Positive Control - one vial of recombinant human IL-17A; lyophilized.	730-01-04	1 vial
Avidin-HRP Conjugate - 120 µL/vial, 100-fold concentrated solution of Avidin conjugate to HRP with preservative.	AVHRP	1 vial
Dilution Buffer – 50 mL of buffered protein based solution with preservative.	DB07	1 bottle
Wash Buffer - 50 mL of 10-fold concentrated buffered surfactant with preservative.	WB01	1 bottle
TMB Substrate Solution - 11 mL of TMB substrate solution.	TMB01	1 bottle
Stop Solution – 11 mL of 0.5M HCl	S-STOP	1 bottle
Plate Sealer	EAPS	1 piece
Plastic Pouch	P01	1 piece

STORAGE

Unopened Kit: Store at 2 - 8° C for up to 6 months. For longer storage, unopened Standard, Positive Control and Detection Antibody Concentrate should be stored at -20° C or -70° C. Do not use kit past expiration date.

Avidin-HRP Conjugate 250-fold concentrated solution and other components may be stored at 2 - 8° C for up to 6 months.

Microplate Wells: Return unused wells to the plastic pouch with the desiccant pack. Microplate may be stored for up to 6 months at 2 - 8° C.

ADDITIONAL MATERIALS REQUIRED

- Microplate reader capable of absorbance measurement at 450 nm.
- Microplate shaker (250 – 300 rpm).
- Microplate washer or manifold dispenser.
- 100 mL and 500 mL graduated cylinders.
- Multi-channel Pipette, Pipettes and pipette tips.
- Deionized or distilled water.

SAMPLE COLLECTION AND STORAGE

Cell Culture Supernates - Remove particulates by centrifugation and assay immediately or aliquot and store samples at $\leq -20^{\circ}\text{C}$. Avoid repeated freeze-thaw cycles.

Serum - Use a serum separator tube (SST) and allow samples to clot for 30 minutes before centrifugation for 15 minutes at 1000 x g. Remove serum and assay immediately or aliquot and store samples at $\leq -20^{\circ}\text{C}$. Avoid repeated freeze-thaw cycles.

Plasma - Collect plasma using EDTA, heparin, or citrate as an anticoagulant. Centrifuge for 15 minutes at 1000 x g within 30 minutes of collection. Assay immediately or aliquot and store samples at $\leq -20^{\circ}\text{C}$. Avoid repeated freeze-thaw cycles.

SAMPLE PREPARATION

Serum and Plasma samples do not require dilutions.

Optimal dilutions should be determined by each laboratory for each application with a pretest. Use polypropylene test tubes.

REAGENT PREPARATION

Bring all reagents to room temperature before use.

Wash Buffer - If crystals have formed in the concentrate, warm to room temperature and mix gently until the crystals have completely dissolved. Dilute 50 mL of Wash Buffer Concentrate into deionized or distilled water (450 mL) to prepare 500 mL of 1x Wash Buffer.

IL-17A Standard - Reconstitute the IL-17A standard with lot specific label of Dilution Buffer. Pipette 250 μL of Dilution Buffer into tubes #1 to #6. Use the stock solution to produce a dilution series (below). Mix each tube thoroughly before the next transfer. The **250 pg/mL** standard serves as the high standard. The Dilution Buffer serves as the zero standard (0 pg/mL).

TUBE	STANDARD	DILUTION BUFFER	CONCENTRATION
Stock	Powder	Lot specific	250 pg/ml
# 1	250 μl of stock	250 μl	125 pg/ml
# 2	250 μl of 1	250 μl	62.5 pg/ml
# 3	250 μl of 2	250 μl	31.25 pg/ml
# 4	250 μl of 3	250 μl	15.6 pg/ml
# 5	250 μl of 4	250 μl	7.8 pg/ml
# 6	250 μl of 5	250 μl	3.9 pg/ml

Positive Control - Reconstitute the Positive Control with lot specific of Dilution Buffer.

Detection Antibody Concentrate - Reconstitute the Detection Antibody Concentrate with lot specific of Dilution Buffer to produce a 100-fold concentrated stock solution. Pipette 10.88 mL of Dilution Buffer into a 15 mL centrifuge tube and transfer 0.12 mL of 10-fold concentrated stock solution to prepare working solution.

Avidin-HRP Conjugate - Pipette 10.88 mL of Dilution Buffer into a 15 mL centrifuge tube and transfer 120 μL of 100-fold concentrated stock solution to prepare working solution. **(protect from light). DO NOT FREEZE.**

ELISA PROTOCOL

Bring all reagents and samples to room temperature before the start of the assay. Blank, standard dilutions, positive control and samples should be assayed in duplicate. ELISA Protocol may need further optimization.

1. Prepare all reagents and working standards as directed in the previous sections.
2. Remove excess microplate strips from the plate frame, return them to the plastic pouch with the desiccant pack.
3. Add 100 μL of **Dilution Buffer** to Blank wells.
4. Add 100 μL of **Standard dilutions (#7 - #1)** in reverse order of serial dilution, **sample**, or **positive control** per well. Cover with plate sealer. Incubate for 2 hours on microplate shaker at room temperature.
5. Aspirate each well and wash, repeating the process three times for a total of four washes. Wash by filling each well with **1x Wash Buffer** (300 μL) using a squirt bottle, manifold dispenser, or autowasher. Let the wash buffer soak for about 1-

2 minutes during each wash step. Complete removal of liquid at each step is essential to good performance. After the last wash, remove any remaining Wash Buffer by aspirating or decanting. Invert the plate and blot it against clean paper towels.

6. Add 100 μ L of **Detection Antibody working solution** to each well. Cover with plate sealer. Incubate for 1 hour on microplate shaker at room temperature.
7. Repeat the aspiration/wash as in step 5.
8. Add 100 μ L of **Avidin-HRP Conjugate working solution** to each well. Incubate for 45 minutes on microplate shaker at room temperature. **Protect from light.**
9. Repeat the aspiration/wash as in step 5, except soak in wash buffer for approximately 1-2 minutes prior to the first aspiration step.
10. Add 100 μ L of **Substrate Solution** to each well. Incubate for 15-20 minutes on microplate shaker at room temperature. **Protect from light.**
11. Add 100 μ L of **Stop Solution** to each well. The color in the wells should change from blue to yellow. If the color in the wells is green, or if the color change does not appear uniform, gently tap the plate to ensure thorough mixing.
12. Determine the optical density of each well within 15 minutes using a microplate reader set to 450 nm.

CALCULATION OF RESULTS

Create a standard curve by plotting the log of the known concentrations of the standard dilutions (x-axis) versus the log of its corresponding O.D. (y-axis) and draw the best fit line through the points. It is recommended to use computer software capable of generating a log-log curve fit to more accurately quantify the standard dilutions.

If samples have been diluted, the concentration read from the standard curve must be multiplied by the dilution factor.

Calculation of samples with a concentration exceeding that of the high standard may result in inaccurate, low human IL-17A levels. Such samples require further external pre-dilution according to expected human IL-17A values with Dilution Buffer in order to precisely quantitate the actual human IL-17A level.

TYPICAL DATA









This standard curve is provided for demonstration only. A standard curve should be generated for each set of samples assayed. The blank is used as a reference for your blank. The O.D. values for the standards have already been corrected (i.e. minus the blank).

STANDARD (PG/ML)	AVERAGE OD450 (CORRECTED)
Blank	0 (0.107)
3.906	0.032
7.813	0.079
15.625	0.155
31.25	0.306
62.5	0.612
125	1.155
250	2.129

SPECIFICITY

PROTEINS	CROSS-REACTIVITY (%)
Human IL-17A	100
Human IL-17F	3
Human IL-12 (p70)	0
Human IL-12 (p40)	0
Human IL-1RN	0
Mouse IL-17A	0
Mouse IL-17F	0

SUMMARY OF ASSAY PROCEDURE

PREPARE REAGENTS, SAMPLES AND STANDARDS

Add 100 µl of standard dilutions, samples, or positive control to each well. Incubate 2 hours on the plate shaker at RT.

Aspirate and wash 4 times.

Add 100 µl Detection Antibody working solution to each well. Incubate 1 hour 30 minutes on the plate shaker at RT.

Aspirate and wash 4 times.

Add 100 µl Avidin-HRP Conjugate working solution to each well. Incubate 45 minutes on the plate shaker at RT. Protect from light.

Aspirate and wash 4 times.

Add 100 µl Substrate Solution to each well. Incubate 3-7 minutes on the plate shaker at RT. Protect from light.

Add 100 µl Stop Solution to each well. Read at 450 nm within 15 min.