

LIPOCALIN-TYPE PROSTAGLANDIN D SYNTHASE (L-PGDS) HUMAN ELISA KIT

FOR THE QUANTITATIVE DETERMINATION OF
HUMAN LPGDS CONCENTRATIONS IN SERUM AND
PLASMA



ALWAYS REFER TO LOT SPECIFIC
PROTOCOL PROVIDED WITH EACH KIT FOR
INSTRUCTIONS. PROTOCOL MUST BE
READ AND CHECK ALL ITEMS OF EACH KIT
BEFORE USING THIS PRODUCT.

FOR RESEARCH USE ONLY. NOT FOR USE IN
DIAGNOSTIC PROCEDURES.

PRODUCT INFORMATION:
THIS KIT IS FOR ONE TIME USE ONLY.

ELISA NAME	LPGDS (HUMAN) ELISA KIT
Catalog No.	SK00025-06
Lot No.	
Formulation	96 T
Standard range	125 – 8000 pg/mL
Sensitivity	30 pg/mL
Sample Volume	100 µL
Sample Type	Serum, Plasma,
Dilution Factor	<i>(Optimal dilutions should be determined by each laboratory for each application)</i>
Specificity	Human LPGDS
Calibration	Mature human LPGDS recombinant
Intra-assay Precision	4 - 6%
Inter-assay Precision	4 - 9%
Storage	2 - 8° C for 6 months, see page 2 for more information
This kit contains sufficient materials to run approximately 35-40 samples duplicated provided that assay is run according to protocol.	

ORDER CONTACT:
AVISCERA BIOSCIENCE, INC.
2348 Walsh Ave., Suite C
Santa Clara, CA 95051
USA
Tel: (408) 982 0300
Email: Sales@AvisceraBioscience.com
Info@AvisceraBioscience.com
www.AvisceraBioscience.net
www.AvisceraBioscience.com

DESCRIPTION

Lipocalin-type prostaglandin D2 synthase (L-PGDS) is a protein that plays a role in metabolism, including glucose transport, fatty liver disease, and insulin resistance. L-PGDS is also involved in the metabolism of arachidonic acid, and is expressed in many tissues, including the brain, heart, liver, and lungs. Lipocalin-type prostaglandin D synthase (L-PGDS) is a protein that may be involved in cardiovascular disease. It may be a predictor of cardiovascular injuries and coronary artery disease.

This LPGDS (Human) ELISA Kit contains the necessary components required for the quantitative measurement of recombinant and/or natural LPGDS from serum and plasma samples in a sandwich ELISA format.

This immunoassay contains recombinant LPGDS and antibodies raised against this protein. Results from this immunoassay have shown to accurately quantify recombinant and natural active LPGDS samples.

ASSAY OVERVIEW

This assay employs the quantitative sandwich enzyme immunoassay technique. The plate is pre-coated with an antibody specific for LPGDS. The capture antibody can bind to the LPGDS in the standard and samples. After washing the plate of any unbound substances, a biotinylated antibody against LPGDS is added to the wells. After another washing of the plate, Streptavidin-HRP Conjugate is added. After the last wash to remove any unbound enzyme, a substrate solution is added to the wells and color develops in direct proportion to the amount of LPGDS bound in the standard solutions or samples. A standard curve can be established and sample values can be read off the standard curve.

PROCEDURAL LIMITATIONS

_FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.

_This ELISA kit should not be used beyond the expiration date on the kit label.

_Do not mix reagents with those from other lots or sources.

_It is important that the Dilution Buffer selected for the standard curve be consistent with the samples being assayed.

_Any modifications in buffers, pipetting technique, washing technique, incubation time or temperature, as well as kit age can cause a change in signal.

_Not all interfering factors have been tested in the immunoassay, therefore the possibility of interference cannot be excluded.

COMPONENTS PROVIDED

DESCRIPTION	CODE	QUANTITY
LPGDS Microplate - 96 well polystyrene microplate (12 strips of 8 wells) coated with an antibody against LPGDS.	025-06-01	1 plate
LPGDS Standard – 8000 pg/vial of rh LPGDS (HEK293 derived) in a buffered protein base with preservative; lyophilized.	025-06-02	1 vial
Detection Antibody Concentrate – 1.2 mL/vial, 10-fold concentrated of biotinylated antibody against LPGDS with preservative; lyophilized.	025-06-03	1 vial
Positive Control - one vial of recombinant LPGDS; lyophilized.	025-06-04	1 vial
Streptavidin-HRP Conjugate - 120 µl/vial, 100-fold concentrated solution of Streptavidin conjugate to HRP.	SAHRP	1 vial
Dilution Buffer – 45 mL of buffered protein based solution with preservative.	DB11C	1 bottle
HRP Diluent Solution – 12 mL of buffered protein based solution with preservative.	DB08B	1 bottle
Wash Buffer 20X - 25 mL of 20-fold concentrated buffered surfactant, with preservative.	WB01	1 bottle
TMB Substrate Solution - 11 mL of TMB substrate solution.	TMB01	1 bottle
Stop Solution - 11 mL of 0.25M HCl solution.	S-STOP	1 bottle
Plate Sealer	EAPS	1 piece
Plastic Pouch	P01	1 piece

STORAGE

Unopened Kit: Store at 2 – 8° C for up to 6 months. For longer storage up to 12 months, unopened Standard, Positive Control, Detection Antibody Concentrate, Dilution Buffer and HRP Diluent Solution should be stored at -20° C. Streptavidin-HRP Conjugate and TMB Substrate Solution should be stored only at 2-8° C. Do not use kit past expiration date.

ADDITIONAL MATERIALS REQUIRED

- Microplate reader capable of absorbance measurement at 450 nm.
- Microplate shaker (250 – 300 rpm).
- Microplate washer or manifold dispenser.
- 100 mL and 500 mL graduated cylinders.
- Multi-channel Pipette, Pipettes and pipette tips.
- Deionized or distilled water.

PRECAUTION

This kit should be handled by those persons who have been trained in and can follow the principles of good laboratory practice. Wear protective clothing such as laboratory overalls, safety glasses and gloves. Care should be taken while handling solutions in this kit to avoid contact with skin or eyes, especially with the stop solution because it contains diluted hydrochloric acid. Wash immediately with water in case of contact on skin or eyes.

SAMPLE COLLECTION AND STORAGE

Serum - Use a serum separator tube (SST) and allow samples to clot for 30 minutes before centrifugation for 15 minutes at 1000 x g. Remove serum and assay immediately or aliquot and store samples at ≤ -20° C. Avoid repeated freeze-thaw cycles.

Plasma - Collect plasma using EDTA, heparin, or citrate as an anticoagulant. Centrifuge for 15 minutes at 1000 x g within 30 minutes of collection. Assay immediately or aliquot and store samples at ≤ -20° C. Avoid repeated freeze-thaw cycles.

Optional: Use Aprotinin (enzyme inhibitor) (Aviscera Bioscience's Order Code: 00700-01-25, 25 TIU for 50 ml sample solution) for ALL sample collection to prevent sample degradation. 0.5 TIU per ml of sample solution.

SAMPLE PREPARATION

Optimal dilutions should be determined by each laboratory for each application with a sample pretest.

Use polypropylene test tubes.

REAGENT PREPARATION

Bring all reagents to room temperature before use.

Wash Buffer - If crystals have formed in the concentrate, warm to room temperature and mix gently until the crystals have completely dissolved. Dilute **25 mL of Wash Buffer Concentrate 20X** into deionized or distilled water (**475 mL**) to prepare 500 mL of 1x Wash Buffer.

LPGDS Standard - Reconstitute the LPGDS standard with 1 mL of Dilution Buffer. This reconstitution produces a stock solution of 8000 pg/mL. Allow the standard to sit for a minimum of 15 minutes with gentle agitation prior to making dilutions. Pipette 300 µL of Dilution Buffer into the tube #1 to #6. Use the stock solution to produce a dilution series (below). Mix each tube thoroughly before the next transfer. The **8000 pg/mL** standard serves as the high standard. The Dilution Buffer serves as the zero standard (0 pg/mL). Store the stock solution of standard at -20 ~ -70 °C for a few days.

TUBE	STANDARD	DILUTION BUFFER	CONCENTRATION
stock	powder	1 ml	8000 pg/ml
# 1	300 µl of stock	300 µl	4000 pg/ml
# 2	300 µl of 1	300 µl	2000 pg/ml
# 3	300 µl of 2	300 µl	1000 pg/ml
# 4	300 µl of 3	300 µl	500 pg/ml
# 5	300 µl of 4	300 µl	250 pg/ml
# 6	300 µl of 5	300 µl	125 pg/ml

Positive Control - Reconstitute the Positive Control with 1 mL of Dilution Buffer. Discard the positive control after use. It is for one time use only.

Detection Antibody - Reconstitute the Detection Antibody Concentrate with 1.2 mL of Dilution Buffer to produce a 10-fold concentrated stock solution. For 96 wells test, freshly pipette 9.45 mL of Dilution Buffer into a 15 ml centrifuge tube and transfer 1.05 mL of 10-fold concentrated stock solution to prepare working solution. *If run a partial strip test, freshly prepare 900 µL per strip (8-wells) of working solution. Store the stock solution of 10-fold*

concentrated detection antibody at $-20\text{ }^{\circ}\text{C}$ for a few days.

Streptavidin-HRP Conjugate – For 96 wells test freshly pipette 11.88 mL of **HRP Diluent solution (DB08B)** into a 15 mL centrifuge tube and transfer 120 μL of 100-fold concentrated stock solution to prepare working solution (**protect from light**).

The working solution of Streptavidin-HRP Conjugate should be freshly prepared and used within 10-20 min. If run a partial strip test, freshly prepare 900 μL per strip (8-wells) of working solution. Store the stock solution of 100-fold concentrated Streptavidin HRP ONLY at $2\text{--}8^{\circ}\text{C}$ for 12 months.

ELISA PROTOCOL

Bring all reagents and samples to room temperature before the start of the assay. Blank, standard dilutions, positive control and samples should be assayed in duplicate. ELISA Protocol may need further optimization.

1. Prepare all reagents and working standards as directed in the previous sections.
2. Add 100 μL per well of Dilution Buffer to Blank wells.
3. Add 100 μL of standard dilutions in reverse order of serial dilution, samples, or positive control per well. Cover with plate sealer. Incubate for 2 hours on microplate shaker at room temperature.
4. Aspirate each well and wash, repeating the process three times for a total of four washes. Wash by filling each well with 1x Wash Buffer (300 μL) using a squirt bottle, manifold dispenser, or autowasher. Complete removal of liquid at each step is essential to good performance. After the last wash, remove any remaining Wash Buffer by aspirating or decanting. Invert the plate and blot it against clean paper towels.
5. Add 100 μL of Detection Antibody working solution to each well. Cover with plate sealer. Incubate for 60 minutes on microplate shaker at room temperature.
6. Repeat the aspiration/wash as in step 4.
7. Add 100 μL of Streptavidin-HRP Conjugate working solution to each well. Incubate for 45 minutes on microplate shaker at room temperature. **Protect from light.**
8. Repeat the aspiration/wash as in step 4.

9. Add 100 μL of Substrate Solution to each well. Incubate for 20 minutes on microplate shaker at room temperature. **Protect from light.**
10. Add 100 μL of Stop Solution to each well. The color in the wells should change from blue to yellow. If the color in the wells is green, or if the color change does not appear uniform, gently tap the plate to ensure thorough mixing.
11. Determine the optical density of each well using a microplate reader set to 450 nm within 5 minutes.

CALCULATION OF RESULTS

Create a standard curve by plotting the log of the known concentrations of the standard dilutions (x-axis) versus the log of its corresponding O.D. (y-axis) and draw the best fit line through the points. It is recommended to use computer software capable of generating a log-log curve or 4-parameter fit to more accurately quantify the standard dilutions.

If samples have been diluted, the concentration read from the standard curve must be multiplied by the dilution factor.

SPECIFICITY

PROTEINS	CROSS-REACTIVITY (%)
Human LPGDS (mature) Rec.	100

Human LPGDS derived from E. Coli or Sf21 may not be detected by this ELISA.

TYPICAL STANDARD CURVE


This standard curve is provided for demonstration only. A new standard curve should be generated for each set of samples assayed.

STANDARD (PG/ML)	AVERAGE OD450NM (CORRECTED)
Blank	0 (0.069)
125	0.039
250	0.081
500	0.160
1000	0.307
2000	0.622
4000	1.201
8000	2.309


- Positive Control: refer to specific lot.

Lipocalin-type Prostaglandin D Synthase (L-PGDS) Human ELISA Kit

Catalog No.: SK00025-06
 Assay Range: 125 ~ 8000 pg/mL
 Sensitivity: 30 pg/mL
 Calibration: rh L-PGDS (HEK293)
 Sample Type: Serum, Plasma, Cell Cultures



Well Position



SUMMARY OF ASSAY PROCEDURE

PREPARE REAGENTS, SAMPLES AND STANDARDS
↓
Add 100 µl of standard dilutions, samples, or positive control to the well. Incubate for 2 hours on the plate shaker at RT.
↓
Aspirate and wash 4 times.
↓
Add 100 µl Detection Antibody working solution to each well. Incubate for 60 minutes on the plate shaker at RT.
↓
Aspirate and wash 4 times.
↓
Add 100 µl Streptavidin-HRP conjugate working solution to each well. Incubate for 45 min on the plate shaker at RT. Protect from light.

↓
Aspirate and wash 4 times.
↓
Add 100 µl Substrate solution to each well. Incubate 20 min on the plate shaker at RT. Protect from light.
↓
Add 100 µl Stop Solution to each well. Read at 450 nm within 5 min.

Aviscera Bioscience Focus on development of Metabolism and Cardiovascular Biomarker ELISA Kits.

Endotrophin Human ELISA Kit	SK00009-08
Nidogen-2 Human ELISA Kit	SK00480-06
EPDR1 ELISA Kit	SK00023-06
LPGDS Human ELISA Kit	SK00025-06
BDNF ELISA Kit	SK00752-01
Lipocalin-13 Human ELISA Kit	SK00648-06
METRNL Human ELISA Kit	SK00478-06
Irisin Human ELISA Kit	SK00170-08
Betatrophin Human ELISA Kit	SK00528-06
CTRP1 Human ELISA Kit	SK00083-06
CTRP15 Human ELISA Kit	SK00393-15
CTRP4 Human ELISA Kit	SK00090-06
FGF21 Human ELISA Kit	SK00145-01