

RAT MCP-1 ELISA KIT

For the quantitative determination of Rat MCP-1 concentrations in serum, plasma and cell culture or tissues.



FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.

PURCHASE INFORMATION:

ELISA Name	Rat MCP-1 ELISA
Catalog No.	SK00220-02
Formulation	96 T
Standard range	31.25-2000 pg/mL
Sensitivity	7 pg/mL
Sample Volume	100 µl
Sample Type	Serum, plasma
Dilution Factor	Optimal dilutions should be determined by each laboratory for each application
Specificity	Rat MCP-1 only
Intra-assay Precision	6-8%
Inter-assay Precision	10%
Storage	2 °C-8 °C

Order Contact:
AVISCIERA BIOSCIENCE INC.
2348 Walsh Ave., Suite C
Santa Clara, CA 95051
Tel: (408) 982 0300
Fax: (408) 982 0301
Email: Sales@AvisceraBioscience.com
Info@AvisceraBioscience.com
www.AvisceraBioscience.com

INTRODUCTION

Rat MCP-1 immunoassay is a 3.5 - 4.5 hour solid phase ELISA designed to measure Rat MCP-1 in cell culture supernates, serum, and plasma. It contains recombinant Rat MCP-1 and antibodies raised against this protein. It has been shown to accurately quantitate recombinant Rat MCP-1. Results obtained with naturally occurring MCP-1 samples showed linear curves that were parallel to the standard curves obtained using the kit standards. These results indicate that the Immunoassay kit can be used to determine relative mass values for natural Rat MCP-1.

PRINCIPLE OF THE ASSAY

This assay employs the quantitative sandwich enzyme immunoassay technique. A monoclonal antibody specific for MCP-1 has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and any MCP-1 present is bound by the immobilized antibody. After washing away any unbound substances, a biotinylated polyclonal antibody specific for MCP-1 is added to the wells. Following a wash to remove any unbound antibody-biotin reagent, HRP link Streptavidin is added to the wells. After washing away any unbound enzyme, a substrate solution is added to the wells and color develops in proportion to the amount of MCP-1 bound in the initial step. The color development is stopped and the intensity of the color is measured.

LIMITATIONS OF THE PROCEDURE

_ FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.

_ The kit should not be used beyond the expiration date on the kit label.

_ Do not mix or substitute reagents with those from other lots or sources.

_ It is important that the Dilution Buffer selected for the standard curve be consistent with the samples being assayed.

_ If samples generate values higher than the highest standard, dilute the samples with the appropriate Dilution Buffer and repeat the assay.

_ Any variation in standard diluent, operator, pipetting technique, washing technique, incubation time or temperature, and kit age can cause variation in binding.

_ This assay is designed to eliminate interference by soluble receptors, binding proteins, and other factors present in biological samples. Until all factors

have been tested in the Immunoassay, the possibility of interference cannot be excluded.

MATERIALS PROVIDED

Description	Code	Quantity
Microplate - 96 well polystyrene microplate (12 strips of 8 wells) coated with a monoclonal antibody against MCP-1.	220-02-01	1 plate
MCP-1 Standard – 2000 pg/vial of recombinant Rat MCP-1 in a buffered protein base with preservatives; lyophilized.	220-02-02	1 vial
Detection Antibody Concentrate – 120 µL / vial, 100-fold concentrated of Biotinylated polyclonal antibody against MCP-1 with preservatives; lyophilized.	220-02-03	1 vial
Positive Control - one of recombinant Rat MCP-1, lyophilized	220-02-04	1 vial
Streptavidin-HRP Conjugate -120 ul/vial, 100-fold concentrated solution of Streptavidin conjugate to HRP with preservatives	SAHRP	1 vial
Dilution Buffer - 60mL/vial of buffered protein based solution with preservatives	DB01	1 vial
Wash Buffer -50 ml/vial, 10-fold concentrated buffered surfactant, with preservative.	WB01	1 vial
TMB Substrate Solution -13 ml / vial of TMB substrate solution	TMB01	1 vial
Stop Solution (0.5 M HCl) , 13 ml /vial of 0.5M HCl	S-STOP	1 vial
Plate Covers – Plate sealer.	EAPS	1

STORAGE

Unopened Kit: Store at 2 - 8° C. Do not use past kit expiration date.

Opened / Reconstituted Reagents: May be stored for up to 1 month at 2 - 8°C.

Standard should be stored for up to 1 month at -70° C.

Microplate Wells: Return unused wells to the foil pouch containing the desiccant pack, reseal along entire edge of zip-seal. May be stored for up to 1 month at 2 - 8° C.

OTHER SUPPLIES REQUIRED

- Microplate reader capable of measuring absorbance at 450 nm, with the correction wavelength set at 540 nm or 570 nm.
- Microplate shaker (250-300rpm).
- Pipettes and pipette tips.
- Deionized or distilled water.
- Squirt bottle, manifold dispenser, or automated microplate washer.
- 100 mL and 500 mL graduated cylinders.

SAMPLE COLLECTION AND STORAGE

Cell Culture Supernates - Remove particulates by centrifugation and assay immediately or aliquot and store samples at ≤-20° C. Avoid repeated freeze-thaw cycles.

Serum - Use a serum separator tube (SST) and allow samples to clot for 30 minutes before centrifugation for 15 minutes at 1000 x g. Remove serum and assay immediately or aliquot and store samples at ≤ -20° C. Avoid repeated freeze-thaw cycles.

Plasma - Collect plasma using EDTA, heparin, or citrate as an anticoagulant. Centrifuge for 15 minutes at 1000 x g within 30 minutes of collection. Assay immediately or aliquot and store samples at ≤-20° C. Avoid repeated freeze-thaw cycles.

REAGENT PREPARATION


Bring all reagents to room temperature before use.

Wash Buffer - If crystals have formed in the concentrate, warm to room temperature and mix gently until the crystals have completely dissolved. Dilute 50 mL of Wash Buffer Concentrate into deionized or distilled water (450 mL) to prepare 500 mL of Wash Buffer.

MCP-1 Standard - Refer to vial label for reconstitution volume. Reconstitute the MCP-1 Standard with 1 ml of Dilution Buffer. This

reconstitution produces a stock solution of 2000 pg/mL. Allow the standard to sit for a minimum of 15 minutes with gentle agitation prior to making dilutions. Pipette 250 µL of the appropriate Dilution Buffer into the tube #1 to #6. Use the stock solution to produce a dilution series (below). Mix each tube thoroughly before the next transfer. The 2000 pg/mL standard serves as the high standard. The appropriate Dilution Buffer serves as the zero standard (0 pg/mL).

Standard	Standard	Reagent Diluent	Concentration
stock	Powder	1000 µl	2000 pg/ml
# 1	250 µl of stock	250 µl	1000 pg/ml
# 2	250 µl of 1	250 µl	500 pg/ml
# 3	250 µl of 2	250 µl	250 pg/ml
# 4	250 µl of 3	250 µl	125 pg/ml
# 5	250 µl of 4	250 µl	62.5 pg/ml
# 6	250 µl of 5	250 µl	31.25 pg/ml



Detection Antibody- Reconstitute the **Detection Antibody concentrated** with 120 µl of Dilution Buffer to produce a 100-fold concentrated stock solution. Pipette 11. 88 mL of the appropriate Dilution Buffer into the 15 ml centrifuge tube and transfer 120 µl of 100-fold concentrated stock solution to prepare working solution.

Streptavidin-HRP Conjugate - Pipette 11. 88 mL of Dilution Buffer into the 15 ml centrifuge tube and transfer 120 µl of 100-fold concentrated stock solution to prepare working solution.

ASSAY PROCEDURE

Bring all reagents and samples to room temperature before use. It is recommended

that standards be assayed in duplicate.

1. Prepare all reagents and working standards as directed in the previous sections.
2. Remove excess micro-plate strips from the plate frame, return them to the foil pouch containing the desiccant pack, reseal.
3. Add 100 μL of Dilution Buffer to Blank well (A1, A2).
4. Add 100 μL of Standard (from B1 to H2, and from H3 to H4), sample, or control per well. Cover with the Sealer. Incubate for 2 hours on micro-plate shaker at room temperature. A plate layout is provided to record standards and samples assayed.
5. Aspirate each well and wash, repeating the process three times for a total of four washes. Wash by filling each well with Wash Buffer (300 μL) using a squirt bottle, manifold dispenser, or autowasher. Complete removal of liquid at each step is essential to good performance. After the last wash, remove any remaining Wash Buffer by aspirating or decanting. Invert the plate and blot it against clean paper towels.
6. Add 100 μL of Detection Antibody working solution to each well. Cover with sealer. Incubate for 2 hours on micro-plate shaker at room temperature.
7. Repeat the aspiration/wash as in step 5.
8. Add 100 μL of **Streptavidin-HRP Conjugate** working solution to each well. Incubate for 40 minutes on micro-plate shaker at room temperature.
9. Repeat the aspiration/wash as in step 5.
10. Add 100 μL of Substrate Solution to each well. Incubate for 20-30 minutes at room temperature.
Protect from light.
11. Add 100 μL of Stop Solution to each well. The color in the wells should change from blue to yellow. If the color in the wells is green or if the color change does not appear uniform, gently tap the plate to ensure thorough mixing.
12. Determine the optical density of each well within 30 minutes, using a micro-plate reader set to 450 nm.

CALCULATION OF RESULTS

Average the duplicate readings for each standard, control, and sample and subtract the average zero standard optical density. Create a standard curve by reducing the data using computer software capable of generating a four parameter logistic (4-PL) curve fit. As an alternative, construct a standard curve by

plotting the mean absorbance for each standard on the y-axis against the concentration on the x-axis and draw a best fit curve through the points on the graph. The data may be linearized by plotting the log of the MCP-1 concentrations versus the log of the O.D. and the best fit line can be determined by regression analysis. This procedure will produce an adequate but less precise fit of the data.

If samples have been diluted, the concentration read from the standard curve must be multiplied by the dilution factor.

TYPICAL DATA

These standard curves are provided for demonstration only. A standard curve should be generated for each set of samples assayed.

CALIBRATION

This immunoassay is calibrated against a highly purified recombinant Rat MCP-1.

SENSITIVITY

Twenty-five assays were evaluated and the minimum detectable dose (MDD) of MCP-1 was 7 pg/mL.

SUMMARY OF ASSAY PROCEDURE

PREPARE REAGENTS, SAMPLES AND STANDARDS
↓
Add 100µl of standard, samples, positive control to each well. Incubate 2 hours on the plate shaker at RT.
↓
Aspirate and wash 4 times.
↓
Add 100 µl Detection Antibody to each well. Incubate 2 hours on the plate shaker at RT.
↓
Aspirate and wash 4 times.
↓
Add 100 µl Streptatvin HRP conjugate to each well. Incubate 40 minutes on the plate shaker at RT.
↓
Aspirate and wash 4 times.
↓
Add 100 µl Substrate to each well. Incubate 20-30min on the bench top. Protect from light.
↓
Add 100 µl Stop Solution to each well. Read 450nm within 30 min

SPECIFICITY

No significant cross-reactivity or interference was observed.

Rat Recombinant Proteins:
MCP-2, IL-6, TNF- α , IL-10