

SOLUBLE NEUREGULIN-4 (NRG4) (HUMAN) ELISA KIT

FOR THE QUANTITATIVE DETERMINATION OF
HUMAN SOLUBLE NRG4 CONCENTRATIONS IN
PLASMA AND SERUM



THIS PROTOCOL OR DATA IS FOR
DEMONSTRATION ONLY. ALWAYS REFER
TO LOT SPECIFIC PROTOCOL PROVIDED
WITH EACH KIT FOR INSTRUCTIONS.
PROTOCOL MUST BE READ BEFORE USING
THIS PRODUCT.

FOR RESEARCH USE ONLY. NOT FOR USE IN
DIAGNOSTIC PROCEDURES.

PRODUCT INFORMATION:

THIS IS FOR ONE TIME USE ONLY.

| | |
|---|--|
| ELISA NAME | HUMAN SOLUBLE NEUREGULIN-4 ELISA |
| Catalog No. | SK00555-04 |
| Lot No.: | |
| Formulation | 96 T |
| Standard range | 250 – 64000 pg/mL |
| Sensitivity | 50 pg/mL |
| Sample Volume | 100 µL |
| Dilution Factor | Optimal dilutions should be determined by each laboratory for each application |
| Sample Type | Serum, Plasma, |
| Specificity | Human |
| Calibration | Human Soluble Neuregulin-4 recombinant |
| Intra-assay Precision | 4 - 6% |
| Inter-assay Precision | 8 - 10% |
| Storage | 2 – 8 °C for 1 month. See page 2 for detail |
| This kit contains sufficient materials to run 35 samples duplicated provided that assay is run according to protocol. | |

Order Contact:
AVISCERA BIOSCIENCE, INC.
 2348 Walsh Ave., Suite C
 Santa Clara, CA 95051
 Tel: (408) 982 0300
 Fax: (408) 982 0301
 Email: Sales@AvisceraBioscience.com
Info@AvisceraBioscience.com
www.AvisceraBioscience.com
www.AvisceraBioscience.net

DESCRIPTION

This Human Soluble Neuregulin-4 (sNRG4) ELISA Kit contains the necessary components required for the quantitative measurement of recombinant and/or natural human soluble NRG4 from serum and plasma in a sandwich ELISA format.

This immunoassay contains recombinant human soluble NRG4 and the pair monoclonal antibodies raised against this protein. Results from this immunoassay have shown to accurately quantify recombinant and natural soluble NRG4 samples.

ASSAY OVERVIEW

This assay employs the quantitative sandwich ELISA format. The plate is pre-coated with a monoclonal antibody specific for human soluble NRG4. The capture antibody can bind to the human soluble NRG4 in the standard and samples. After washing the plate of any unbound substances, a biotinylated monoclonal antibody against human soluble NRG4 is added to the wells. After another washing of the plate, the Streptavidin-HRP Conjugate is added. After the last wash to remove any unbound enzyme, a substrate solution (TMB) is added to the wells and color develops in direct proportion to the amount of human soluble NRG4 bound in the standard solutions or samples. A standard curve can be established and sample values can be read off the standard curve.

PROCEDURAL LIMITATIONS

_FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.

_This ELISA kit should not be used beyond the expiration date on the kit label.

_Do not mix reagents with those from other lots or sources.

_It is important that the Dilution Buffer selected for the standard curve be consistent with the samples being assayed.

_Each laboratory must determine the optimal dilution factors for the samples being assayed with a pretest. If samples generate values that are not within the dynamic range of the standard curve, further concentrate or dilute the samples as required with Dilution Buffer and repeat the assay.

_Any modifications in buffers, pipetting technique, washing technique, incubation time or temperature, as well as kit age can cause a change in signal.

_Not all interfering factors have been tested in the immunoassay, therefore the possibility of interference cannot be excluded.

COMPONENTS PROVIDED

| DESCRIPTION | CODE | QUANTITY |
|---|------------------|-----------------|
| NRG4 Microplate – 96 well microplate coated with antibody specific for human soluble NRG4. | 555-04-01 | 1 plate |
| soluble NRG4 Standard – refer to lot of lyophilized recombinant soluble NRG4. | 555-04-02 | 1 vial |
| Detection Antibody Concentrate – refer to lot of lyophilized biotinylated antibody against human Soluble NRG4. | 555-04-03 | 1 vial |
| Positive Control – one vial of lyophilized recombinant human soluble NRG4 | 555-04-04 | 1 vial |
| Streptavidin-HRP Conjugate – 120 µL/vial of 100-fold concentrated solution of Streptavidin-HRP conjugate. | SAHRP | 1 vial |
| Dilution Buffer – 40 mL of buffered solution with preservative. | DB03 | 1 bottle |
| Antibody & HRP Diluent Solution – 25 mL of buffered solution with preservative. | DB11 | 1 bottle |
| Wash Buffer – 50 mL of 10-fold concentrated buffered surfactant with preservative. | WB01 | 1 bottle |
| TMB Substrate Solution – 11 mL of TMB substrate solution. | TMB01 | 1 bottle |
| Stop Solution – 11 mL of 0.5M HCl. | S-STOP | 1 bottle |
| Plate Sealer | EAPS | 1 piece |
| Plastic Pouch | P01 | 1 piece |

STORAGE

Unopened Kit: Store at 2 – 8 °C for up to 1 month. For longer storage for up to 10 months, unopen Standard, Positive Control, Detection Antibody Concentrate, Dilution Buffer and Antibody & RP Diluent Solution should be stored at -20 °C. Streptavidin HRP Conjugate and TMB Substrate Solution should be stored only at 2 – 8 °C. Do not use kit past expiration date.

ADDITIONAL MATERIALS REQUIRED

- Microplate reader capable of absorbance measurement at 450 nm.
- Microplate shaker (250 – 300 rpm).
- Microplate washer or manifold dispenser.
- 100 mL and 500 mL graduated cylinders.
- Multi-channel Pipette, Pipettes and pipette tips.
- Deionized or distilled water.

PRECAUTION

This kit should be handled by those persons who have been trained in and can follow the principles of good laboratory practice. Wear protective clothing such as laboratory overalls, safety glasses and gloves. Care should be taken while handling solutions in this kit to avoid contact with skin or eyes, especially with the stop solution because it contains diluted hydrochloric acid. Wash immediately with water in case of contact on skin or eyes.

SAMPLE COLLECTION AND STORAGE

Serum - Use a serum separator tube (SST) and allow samples to clot for 30 minutes before centrifugation for 15 minutes at 1000 x g. Remove serum and assay immediately or aliquot and store samples at $\leq -20^{\circ}\text{C}$. Avoid repeated freeze-thaw cycles.

Plasma - Collect plasma using EDTA, heparin, or citrate as an anticoagulant. Centrifuge for 15 minutes at 1000 x g within 30 minutes of collection. Assay immediately or aliquot and store samples at $\leq -20^{\circ}\text{C}$. Avoid repeated freeze-thaw cycles.

Cell Culture - All samples should be used animal free media. The fetal bovine serum or other animal serum may cross-react and or interface with this elisa assay.

Note: Use Aprotinin (enzyme inhibitor) (Code No.: 00700-01-25) for ALL sample collection to prevent sample degradation. 0.5 TIU per ml of sample solution.

SAMPLE PREPARATION

Serum samples may need 10-40 fold dilution.

Optimal dilutions should be determined by each laboratory for each application with a pretest.

Use polypropylene test tubes.

REAGENT PREPARATION

Bring all reagents to room temperature before use.

Wash Buffer – Dilute 50 mL of Wash Buffer Concentrate into 450 mL distilled or deionized water to make 500 mL of 1x Wash Buffer. If crystals have formed in the concentrate, warm bottle in a water bath until the crystals have completely dissolved.

NRG4 Standard – Reconstitute the human soluble NRG4 standard with refer to lot of Dilution Buffer. The concentration of the reconstituted stock solution is refer to lot. Allow the stock standard to sit for at least 15 minutes with gentle agitation until completely dissolved prior to making standard dilutions (see below). Mix each tube thoroughly before the next transfer.

| TUBE | STANDARD | DILUTION BUFFER | CONCENTRATION |
|-------|----------------------------|-------------------|---------------|
| Stock | powder | Refer to lot | 64 ng/mL |
| # 1 | 250 μL of stock | 250 μL | 32 ng/mL |
| # 2 | 250 μL of 1 | 250 μL | 16 ng/mL |
| # 3 | 250 μL of 2 | 250 μL | 8 ng/mL |
| # 4 | 250 μL of 3 | 250 μL | 4 ng/mL |
| # 5 | 250 μL of 4 | 250 μL | 2 ng/mL |
| # 6 | 250 μL of 5 | 250 μL | 1 ng/mL |
| # 7 | 250 μL of 6 | 250 μL | 0.5 ng/mL |
| # 8 | 250 μL of 7 | 250 μL | 0.25 ng/mL |

Detection Antibody - Reconstitute the Detection Antibody Concentrate with refer to lot of Antibody & HRP Diluent Solution to produce a 10-fold concentrated stock solution. Allow the concentrated solution to sit for at least 5 minutes until completely dissolved. Pipette refer to lot of Antibody & HRP Diluent Solution into a 15 mL centrifuge tube and transfer refer to lot of 10-fold concentrated stock solution to prepare working solution.

Streptavidin-HRP Conjugate - Pipette 11.88 mL of Antibody & HRP Diluent Solution into a 15 mL centrifuge tube and transfer 120 μL of 100-fold concentrated stock solution to prepare working solution. **Note: (protect from light). DO NOT FREEZE.**

Positive Control - Reconstitute the Positive Control with refer to lot of Dilution Buffer.

ELISA PROTOCOL

Bring all reagents and samples to room temperature before the start of the assay. Blank, standard dilutions, positive control and samples should be assayed in duplicate. ELISA Protocol may need further optimization.

1. Prepare all reagents, standard dilutions, positive control and samples as directed previously.
2. Remove unneeded microplate strips from the plate frame and return them to the plastic pouch with the desiccant pack.
3. Add 100 μ L per well of **Dilution Buffer** to Blank wells.
4. Add 100 μ L per well of **Standard Dilutions** in reverse order of serial dilution, **sample**, or **positive control**. Cover with plate sealer and incubate for 2 hours on microplate shaker at room temperature. (Please see plate layout provided.)
5. Aspirate and wash each well with 300 μ L of **1x Wash Buffer** four times. After the last wash, aspirate any remaining 1x Wash Buffer, invert the plate and blot against clean paper towel(s).
6. Add 100 μ L per well of **Detection Antibody working solution**. Cover with plate sealer and incubate for 2 hours on microplate shaker at room temperature.
7. Repeat the aspiration and wash as in step 5.
8. Add 100 μ L per well of working solution of **Streptavidin HRP conjugate**. Cover with plate sealer and incubate for 60 minutes on microplate shaker at room temperature. **Protect from light.**
9. Repeat the aspiration and wash as in step 5.
10. Add 100 μ L per well of **Substrate Solution**. Incubate for refer to lot on microplate shaker at room temperature. **Protect from light.**
11. Add 100 μ L per well of **Stop Solution**. The color in the wells should change from blue to yellow. If the color change does not appear uniform, gently tap the plate to ensure thorough mixing.
12. Read plate using a microplate reader set to 450 nm within 15 minutes.

CALCULATION OF RESULTS

Create a standard curve by plotting the log of the known concentrations of the standard dilutions (x-axis) versus the log of its corresponding O.D. (y-axis) and draw the best fit line through the points. It is recommended to use computer software capable of generating a log-log curve fit to more accurately quantify the standard dilutions.

If samples have been diluted, the concentration read from the standard curve must be multiplied by the dilution factor.

SPECIFICITY









| PROTEIN | CROSS-REACTIVITY |
|--------------------|------------------|
| Human Soluble NRG4 | 100% |
| Human Soluble NRG1 | 0 |
| Human BDNF | 0 |
| Human Pro BDNF | 0 |

TYPICAL STANDARD CURVE

This standard curve is for demonstration only. A new standard curve should be made for each set of samples assayed.

| STANDARD (NG/ML) | AVERAGE OD450 NM (CORRECTED) |
|------------------|------------------------------|
| Blank | 0 (refer to lot) |
| 0.25 | 0.031 |
| 0.5 | 0.069 |
| 1 | 0.118 |
| 2 | 0.217 |
| 4 | 0.399 |
| 8 | 0.772 |
| 16 | 1.443 |
| 32 | 2.239 |
| 64 | 2.832 |

SUMMARY OF ASSAY PROCEDURE

| |
|---|
| PREPARE REAGENTS, SAMPLES AND STANDARD DILUTIONS |
|  |
| Add 100 μ L of standard dilutions, samples and positive control. Cover with plate sealer and incubate 2 hours on microplate shaker at RT. |
|  |
| Aspirate and wash 4 times. |
|  |
| Add 100 μ L per well of Detection Antibody working solution. Cover with plate sealer and incubate 2 hours on microplate shaker at RT. |
|  |
| Aspirate and wash 4 times. |
|  |
| Add 100 μ L per well of Streptavidin-HRP Conjugate working solution. Cover with plate sealer and incubate 60 minutes on microplate shaker at RT. Protect from light. |
|  |
| Aspirate and wash 4 times. |
|  |
| Add 100 μ L per well of Substrate Solution. Incubate refer to lot on microplate shaker at RT. Protect from light. |
|  |
| Add 100 μ L per well of Stop Solution. Read at 450 nm within 5 minutes. |