

# HUMAN PROSTASIN ELISA KIT

FOR THE QUANTITATIVE DETERMINATION  
OF HUMAN PROSTASIN CONCENTRATIONS  
IN SERUM AND PLASMA



## PRODUCT INFORMATION:

ELISA NAME	HUMAN PROSTASIN ELISA
Catalog No.	SK00156-06
Lot No.	
Formulation	96 T
Standard Range	39 - 2500 pg/ml
Sensitivity	10 pg/ml
Sample Volume	100 µl per well
Sample Type	Serum, EDTA Plasma
Specificity	Human Prostasin
Calibration	Human Prostasin Recombinant
Dilution Factor	<b>Optimal dilutions should be determined by each laboratory for each application</b>
Intra-assay Precision	6 - 8%
Inter-assay Precision	8 - 12%
Storage	2 – 8° C
This kit contains sufficient materials to run 35 samples duplicated provided that assay is run according to protocol.	

ALWAYS REFER TO LOT SPECIFIC PROTOCOL PROVIDED WITH EACH KIT FOR INSTRUCTIONS. PROTOCOL MUST BE READ BEFORE USING THIS PRODUCT.

FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.

**ORDER CONTACT:**  
**AVISCIERA BIOSCIENCE, INC.**  
 2348 WALSH AVE., SUITE C  
 SANTA CLARA, CA 95051  
 USA  
 TEL: (408) 982 0300  
 Email: [Info@AvisceraBioscience.com](mailto:Info@AvisceraBioscience.com)  
 Website: [www.AvisceraBioscience.com](http://www.AvisceraBioscience.com)

## DESCRIPTION

This Human Prostasin ELISA Kit contains the necessary components required for the quantitative measurement of recombinant and/or natural human Prostasin from serum and plasma in a sandwich ELISA format.

This immunoassay contains recombinant human Prostasin and antibodies raised against this protein. Results from this immunoassay have shown to accurately quantify recombinant and natural Prostasin samples.

## ASSAY OVERVIEW

This assay employs the quantitative sandwich ELISA format. The plate is pre-coated with an antibody specific for human Prostasin. The capture antibody can bind to the human Prostasin in the standard and samples. After washing the plate of any unbound substances, a biotinylated antibody against human Prostasin is added to the wells. After another washing of the plate, Streptavidin-HRP Conjugate is added. After the last wash to remove any unbound enzyme, a substrate solution (TMB) is added to the wells and color develops in direct proportion to the amount of human Prostasin bound in the standard solutions or samples. A standard curve can be established and sample values can be read off the standard curve.

## PROCEDURAL LIMITATIONS

\_FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.

\_This ELISA kit should not be used beyond the expiration date on the kit label.

\_Do not mix reagents with those from other lots or sources.

\_It is important that the Dilution Buffer selected for the standard curve be consistent with the samples being assayed.

\_Each laboratory must determine the optimal dilution factors for the samples being assayed with a pretest. If samples generate values that are not within the dynamic range of the standard curve, further concentrate or dilute the samples as required with Dilution Buffer and repeat the assay.

\_Any modifications in buffers, pipetting technique, washing technique, incubation time or temperature, as well as kit age can cause a change in signal.

\_Not all interfering factors have been tested in the immunoassay, therefore the possibility of interference cannot be excluded.

## COMPONENTS PROVIDED

DESCRIPTION	CODE	QUANTITY
<b>Prostasin Microplate</b> - 96 well polystyrene microplate (12 strips of 8 wells) coated with IgG against human Prostasin.	<b>156-06-01</b>	<b>1 plate</b>
<b>Prostasin Standard</b> – 2500 pg/vial of CTRP15 in a buffered protein base with preservative; lyophilized.	<b>156-06-02</b>	<b>1 vial</b>
<b>Detection Antibody Concentrate</b> – 1.2mL/vial, 10-fold concentrate of biotinylated IgG against human Prostasin with preservative; lyophilized.	<b>156-06-03</b>	<b>1 vial</b>
<b>Positive Control</b> - one vial of human prostasin; lyophilized.	<b>156-06-04</b>	<b>1 vial</b>
<b>Streptavidin-HRP Conjugate</b> - 120 µL of 100-fold concentrated Streptavidin-HRP Conjugate.	<b>SAHRP</b>	<b>1 vial</b>
<b>Dilution Buffer</b> - 60 mL of buffered protein based solution with preservative.	<b>DB08</b>	<b>1 bottle</b>
<b>HRP Diluent Solution</b> - 12 mL of buffered protein based solution with preservative.	<b>DB06</b>	<b>1 bottle</b>
<b>Wash Buffer</b> - 50 mL of 10-fold concentrated buffered surfactant, with preservative.	<b>WB01</b>	<b>1 bottle</b>
<b>TMB Substrate Solution</b> - 11 mL of TMB substrate solution.	<b>TMB01</b>	<b>1 bottle</b>
<b>Stop Solution</b> - 11 mL of 0.5M HCl.	<b>S-STOP</b>	<b>1 bottle</b>
<b>Plate Sealer</b>	<b>EAPS</b>	<b>1 piece</b>
<b>Plastic Pouch</b>	<b>P01</b>	<b>1 piece</b>

## STORAGE

**Unopened Kit:** Store at 2 – 8° C for up to 8 months. For longer storage, unopened Standard, Positive Control and Detection Antibody Concentrate should be stored at -20° C or -70° C. Do not use kit past expiration date.

**Opened / Reconstituted Reagents:** Reconstituted Standard (stock) solution and Detection Antibody concentrated solution SHOULD BE STORED at -20° C or -70° C for up to one month. Streptavidin-HRP

Conjugate 100-fold concentrated solution and TMB Substrate Solution can be stored at 2 – 8° C for up to 8 months (**DO NOT FREEZE** and **PROTECT FROM LIGHT**). All other components may be stored at 2 – 8° C for up to 8 months.

**Microplate Wells:** Return unused strips to the plastic pouch with the desiccant pack. Microplate may be stored for up to 6 months at 2 – 8° C after opening.

### ADDITIONAL MATERIALS REQUIRED

- Microplate reader capable of absorbance measurement at 450 nm.
- Microplate shaker (250 – 300 rpm).
- Microplate washer or manifold dispenser.
- 100 mL and 500 mL graduated cylinders.
- Multi-channel Pipette, Pipettes and pipette tips.
- Deionized or distilled water.

### PRECAUTION

This kit should be handled by those persons who have been trained in and can follow the principles of good laboratory practice. Wear protective clothing such as laboratory overalls, safety glasses and gloves. Care should be taken while handling solutions in this kit to avoid contact with skin or eyes, especially with the stop solution because it contains diluted hydrochloric acid. Wash immediately with water in case of contact on skin or eyes.

### SAMPLE COLLECTION AND STORAGE

**Serum** - Use a serum separator tube (SST) and allow samples to clot for 30 minutes before centrifugation for 15 minutes at 1000 x g. Remove serum and assay immediately or aliquot and store samples at ≤ -20° C. Avoid repeated freeze-thaw cycles.

**Plasma** - Collect plasma using EDTA, heparin, or citrate as an anticoagulant. Centrifuge for 15 minutes at 1000 x g within 30 minutes of collection. Assay immediately or aliquot and store samples at ≤ -20° C. Avoid repeated freeze-thaw cycles.

**Optional: Use Aprotinin (enzyme inhibitor) for ALL sample collection to prevent sample degradation. 0.5 TIU per ml of sample solution.**

### SAMPLE PREPARATION

Serum and plasma samples may not require dilution.

**Optimal dilutions should be determined by each laboratory for each application.**

**Use polypropylene test tubes.**

### REAGENT PREPARATION

**Bring all reagents to room temperature before use.**

**Wash Buffer** - If crystals have formed in the concentrate, warm to room temperature and mix gently until the crystals have completely dissolved. Dilute 50 mL of Wash Buffer Concentrate into deionized or distilled water (450 mL) to prepare 500 mL of 1x Wash Buffer.

**Prostasin Standard** - Reconstitute the Prostasin standard with 1 mL of Dilution Buffer. This reconstitution produces a stock solution of 40 ng/mL. Allow the standard to sit for a minimum of 15 minutes with gentle agitation prior to making dilutions. Pipette 250 µL of Dilution Buffer into tubes #1 to #6. Use the stock solution to produce a dilution series (below). Mix each tube thoroughly before the next transfer. The **2500 pg/mL** standard serves as the high standard. The Dilution Buffer serves as the zero standard (0 ng/mL).

TUBE	STANDARD	DILUTION BUFFER	CONCENTRATION
Stock	Powder	1000 µl	2500 pg/ml
# 1	250 µl of stock	250 µl	1250 pg/ml
# 2	250 µl of 1	250 µl	62.5 pg/ml
# 3	250 µl of 2	250 µl	31.25 pg/ml
# 4	250 µl of 3	250 µl	15.6 pg/ml
# 5	250 µl of 4	250 µl	7.8 pg/ml
# 6	250 µl of 5	250 µl	3.9 pg/ml

**Positive Control** – Reconstitute the Positive Control with 1 mL of Dilution Buffer to prepare working solution. **Note:** Positive Control solution could be reused within a few days if stored at -20° C or -70° C.

**Detection Antibody Concentrate** – Reconstitute the Detection Antibody Concentrate with 1.2 mL of Dilution Buffer to produce a 10-fold concentrated stock solution. Pipette 10.8 mL of Dilution Buffer into a 15 mL centrifuge tube and transfer 1.2 mL of 10-fold concentrated stock solution to prepare working solution.

**Streptavidin HRP Conjugate** – Pipette 11.88 mL of **HRP Diluent Solution (DB06)** into a 15 mL centrifuge tube and transfer 120 µL of 100-fold concentrated

stock solution to prepare working solution. **Note:** 1x working solution of Streptavidin-HRP Conjugate should be used within a few days (**protect from light**).

## ELISA PROTOCOL

**Bring all reagents and samples to room temperature before the start of the assay. Blank, standard dilutions, positive control and samples should be assayed in duplicate. ELISA Protocol may need further optimization.**

1. Prepare all reagents and working standards as directed in the previous sections.
2. Remove excess microplate strips from the plate frame, return them to the plastic pouch with the desiccant pack.
3. Add 100  $\mu\text{L}$  per well of Dilution Buffer to Blank wells.
4. Add 100  $\mu\text{L}$  of Standard dilutions, sample, or positive control per well. Cover with plate sealer. Incubate for 2 hours on microplate shaker at room temperature.
5. Aspirate each well and wash, repeating the process three times for a total of four washes. Wash by filling each well with 1x Wash Buffer (300 $\mu\text{L}$ ) using a squirt bottle, manifold dispenser, or autowasher. Complete removal of liquid at each step is essential to good performance. After the last wash, remove any remaining Wash Buffer by aspirating or decanting. Invert the plate and blot it against clean paper towels.
6. Add 100  $\mu\text{L}$  of Detection Antibody working solution to each well. Cover with plate sealer. Incubate for 2 hours on microplate shaker at room temperature.
7. Repeat the aspiration/wash as in step 5.
8. Add 100  $\mu\text{L}$  of Streptavidin-HRP working solution to each well. Cover with plate sealer. Incubate for 45 minutes on microplate shaker at room temperature. **Protect from light.**
9. Repeat the aspiration/wash as in step 5.
10. Add 100  $\mu\text{L}$  of TMB Substrate Solution to each well. Incubate for 3-5 minutes on microplate shaker at room temperature. **Protect from light.**
11. Add 100  $\mu\text{L}$  of Stop Solution to each well. The color in the wells should change from blue to yellow. If the color in the wells is green, or if the color change does not appear uniform, gently tap the plate to ensure thorough mixing.
12. Determine the optical density of each well within 15 minutes, using a microplate reader set to 450 nm.

## CALCULATION OF RESULTS

Create a standard curve by plotting the log of the known concentrations of the standard dilutions (x-axis) versus the log of its corresponding O.D. (y-axis) and draw the best fit line through the points. It is recommended to use computer software capable of generating a log-log curve fit to more accurately quantify the standard dilutions.

If samples have been diluted, the concentration read from the standard curve must be multiplied by the dilution factor.

## TYPICAL STANDARD CURVE

This standard curve is provided for demonstration only. A new standard curve should be generated for each set of samples assayed.









STANDARD (PG/ML)	CORRECTED (450NM)
Blank	0 (0.105)
39	0.044
78	0.102
156.25	0.201
312.5	0.367
625	0.627
1250	1.034
2500	1.517

## SPECIFICITY

PROTEINS	CROSS-REACTIVITY (%)
Human Prostasin	100
Human NSE	0
Human sHer2	0
Human Pro GRP	0
Human AGR2	0

---

**SUMMARY OF ASSAY PROCEDURE**

<b>PREPARE REAGENTS, SAMPLES AND STANDARDS</b>

Add 100 µl of standard dilutions, samples, or positive control to the well. Incubate 2 hours on the plate shaker at RT.

Aspirate and wash 4 times.

Add 100 µl of Detection Antibody working solution to each well. Incubate 2 hours on the plate shaker at RT.

Aspirate and wash 4 times.

Add 100 µl of Streptavidin-HRP working solution to each well. Incubate 45 minutes on the plate shaker at RT. <b>Protect from light.</b>

Aspirate and wash 4 times.

Add 100 µl of TMB Substrate Solution to each well. Incubate 3-5 min on the plate shaker at RT. <b>Protect from light.</b>

Add 100 µl of Stop Solution to each well. Read 450nm within 15 min.