

HUMAN SECRETED PROTEIN ACIDIC AND RICH IN CYSTEINE (SPARC) /OSTEONECTIN ELISA KIT

FOR THE QUANTITATIVE DETERMINATION
OF HUMAN SPARC CONCENTRATIONS IN
SERUM



THIS IS FOR DEMONSTRATION ONLY.
ALWAYS REFER TO LOT SPECIFIC PROTOCOL
PROVIDED WITH EACH KIT FOR
INSTRUCTIONS. PROTOCOL MUST BE
READ BEFORE USING THIS PRODUCT.

FOR RESEARCH USE ONLY. NOT FOR USE IN
DIAGNOSTIC PROCEDURES.

PRODUCT INFORMATION:

NAME	HUMAN SPARC ELISA KIT
Catalog No.	SK00766-08
Formulation	96 T
Lot No.	
Standard range	0.39-25 ng/mL
Sensitivity	0.15 ng/mL
Sample Volume	100 μ L
Dilution Factor	40 for serum (Optimal dilutions should be determined by each laboratory for each application)
Sample Type	Serum
Specificity	Human SPARC
Calibration	Human SPARC recombinant (HEK293 derived)
Intra-assay Precision	4 - 6%
Inter-assay Precision	8 - 10%
Storage	2 – 8 °C
This kit contains sufficient materials to run 35 samples duplicated provided that assay is run according to protocol.	

Order Contact:
AVISCIERA BIOSCIENCE, INC.
2348 Walsh Ave., Suite C
Santa Clara, CA 95051
USA

Tel: (408) 982 0300

Fax: (408) 982 0301

Email: Sales@AvisceraBioscience.com

Info@AvisceraBioscience.com

www.AvisceraBioscience.com

DESCRIPTION

This Human SPARC ELISA Kit contains the necessary components required for the quantitative measurement of recombinant and/or natural human SPARC from serum in a sandwich ELISA format.

This immunoassay contains recombinant human SPARC and antibodies raised against this protein. Results from this immunoassay have shown to accurately quantify recombinant and natural SPARC samples.

ASSAY OVERVIEW

This assay employs the quantitative sandwich ELISA format. The plate is pre-coated with an antibody specific for human SPARC. The capture antibody can bind to the human SPARC in the standard and samples. After washing the plate of any unbound substances, an antibody against human SPARC is added to the wells. After another washing of the plate, Anti rabbit IgG-HRP Conjugate is added. After the last wash to remove any unbound enzyme, a substrate solution (TMB) is added to the wells and color develops in direct proportion to the amount of human SPARC bound in the standard solutions or samples. A standard curve can be established and sample values can be read off the standard curve.

PROCEDURAL LIMITATIONS

_FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.

_This ELISA kit should not be used beyond the expiration date on the kit label.

_Do not mix reagents with those from other lots or sources.

_It is important that the Dilution Buffer selected for the standard curve be consistent with the samples being assayed.

_Each laboratory must determine the optimal dilution factors for the samples being assayed with a pretest. If samples generate values that are not within the dynamic range of the standard curve, further concentrate or dilute the samples as required with Dilution Buffer and repeat the assay.

_Any modifications in buffers, pipetting technique, washing technique, incubation time or temperature, as well as kit age can cause a change in signal.

_Not all interfering factors have been tested in the immunoassay, therefore the possibility of interference cannot be excluded.

COMPONENTS PROVIDED

DESCRIPTION	CODE	QUANTITY
SPARC Microplate – 96 well microplate coated with an antibody specific for human SPARC.	766-08-01	1 plate
SPARC Standard – lot specific of lyophilized recombinant human SPARC.	766-08-02	1 vial
Detection Antibody Concentrate – lot specific of 10-fold concentrate of lyophilized antibody against human SPARC.	766-08-03	1 vial
Positive Control – one vial of lyophilized recombinant human SPARC.	766-08-04	1 vial
Anti Rabbit IgG-HRP Conjugate – 120 µL/vial of 100-fold concentrated solution of Anti Rabbit IgG-HRP conjugate.	ARIGHRP	1 vial
Dilution Buffer – 50 mL of buffered solution with preservative.	DB06	1 bottle
Antibody and HRP Diluent Solution – 30 ml of buffered protein based solution with preservative.	DB08B	1 bottle
Wash Buffer – 50 mL of 10-fold concentrated buffered surfactant with preservative.	WB01	1 bottle
TMB Substrate Solution – 11 mL of TMB substrate solution.	TMB01	1 bottle
Stop Solution – 11 mL of 0.5M HCl.	S-STOP	1 bottle
Plate Sealer	EAPS	1 piece
Plastic Pouch	P01	1 piece

STORAGE

Unopened Kit: Store at 2 – 8° C for up to 12 months. For longer storage, unopened Standard, Positive Control and Detection Antibody Concentrate should be stored at -20° C or -70° C. Do not use kit past expiration date.

Microplate Wells: Return unused strips to the plastic pouch with the desiccant pack. Microplate may be stored for up to 6 months at 2 – 8° C after opening.

ADDITIONAL MATERIALS REQUIRED

- Microplate reader capable of absorbance measurement at 450 nm.
- Microplate shaker (250 – 300 rpm).
- Microplate washer or manifold dispenser.
- 100 mL and 500 mL graduated cylinders.
- Multi-channel Pipette, Pipettes and pipette tips.
- Deionized or distilled water.

PRECAUTION

This kit should be handled by those persons who have been trained in and can follow the principles of good laboratory practice. Wear protective clothing such as laboratory overalls, safety glasses and gloves. Care should be taken while handling solutions in this kit to avoid contact with skin or eyes, especially with the stop solution because it contains diluted hydrochloric acid. Wash immediately with water in case of contact on skin or eyes.

SAMPLE COLLECTION AND STORAGE

Serum – Use a serum separator tube (SST). Allow blood to clot for 30 minutes. Centrifuge at 1000 x g for 15 minutes and collect serum. Assay samples immediately or aliquot and store at $\leq -20^{\circ}$ C. Avoid repeated freeze-thaw cycles.

Optional: Use Aprotinin (enzyme inhibitor) for ALL sample collection to prevent sample degradation. 0.5 TIU per mL of sample solution.

SAMPLE PREPARATION

Serum samples may require a 40-fold dilution. A suggested 40-fold dilution is 10 μ L sample + 390 μ L Dilution Buffer.

Optimal dilutions should be determined by each laboratory for each application with a pretest. Use polypropylene test tubes.

REAGENT PREPARATION

Bring all reagents to room temperature before use.

Wash Buffer – If crystals have formed in the concentrate, warm bottle in a water bath until the crystals have completely dissolved. Dilute 50 mL of Wash Buffer Concentrate into 450 mL distilled or deionized water to make 500 mL of 1x Wash Buffer.

SPARC Standard – Reconstitute the SPARC standard with lot specific of Dilution Buffer. Allow the standard to sit for a minimum of 15 minutes with gentle agitation prior to making dilutions.

Pipette 250 μ L of Dilution Buffer into the tube #2 to #7. Use the stock solution to produce a dilution series (below). Mix each tube thoroughly before the next transfer. The **12.5 ng/mL** standard serves as the high standard. The Dilution Buffer serves as the zero standard (0 ng/mL).

TUBE	STANDARD	DILUTION BUFFER	CONCENTRATION
Stock	powder	Lot specific	
# 1	Lot specific	Lot specific	12.5 ng/mL
# 2	250 μ L of 1	250 μ L	6.25 ng/mL
# 3	250 μ L of 2	250 μ L	3.125 ng/mL
# 4	250 μ L of 3	250 μ L	1.563 ng/mL
# 5	250 μ L of 4	250 μ L	0.78 ng/mL
# 6	250 μ L of 5	250 μ L	0.39 ng/mL
# 7	250 μ L of 6	250 μ L	0.19 ng/mL

Positive Control - Reconstitute the Positive Control with lot specific of Dilution Buffer.

Detection Antibody - Reconstitute the Detection Antibody Concentrate with lot specific of **Antibody and HRP Diluent Solution (DB08B)** to produce a 10-fold concentrated stock solution. Allow the concentrated solution to sit for at least 5 minutes until completely dissolved. Pipette 9.45 mL of **Antibody and HRP Diluent Solution (DB08B)** into a 15 mL centrifuge tube and transfer 1.05 mL of 10-fold concentrated stock solution to prepare working solution.

Anti Rabbit IgG-HRP Conjugate - Pipette 11.88 mL of **Antibody and HRP Diluent Solution (DB08B)** into a 15 mL centrifuge tube and transfer 120 μ L of 100-fold concentrated stock solution to prepare working solution. **(protect from light). DO NOT FREEZE.**

ELISA PROTOCOL

Bring all reagents and samples to room temperature before the start of the assay. Blank, standard dilutions, positive control and samples should be assayed in duplicate. ELISA Protocol may need further optimization.

1. Prepare all reagents, standard dilutions, positive control and samples as directed previously.
2. Remove unneeded microplate strips from the plate frame and return them to the plastic pouch with the desiccant pack.

3. Add 100 µL per well of **Dilution Buffer** to Blank wells.
4. Add 100 µL per well of **Standard dilutions, samples, or positive control**. Cover with plate sealer and incubate for 2 hours on microplate shaker at room temperature.
5. Aspirate and wash each well with 300 µL of **1x Wash Buffer** four times. After the last wash, aspirate any remaining 1x Wash Buffer, invert the plate and blot against clean paper towel(s).
6. Add 100 µL per well of **Detection Antibody working solution**. Cover with plate sealer and incubate for 2 hours on microplate shaker at room temperature.
7. Repeat the aspiration and wash as in step 5.
8. Add 100 µL per well of **ARIGHRP Conjugate working solution**. Cover with plate sealer and incubate for 60 minutes on microplate shaker at room temperature. **Protect from light.**
9. Repeat the aspiration and wash as in step 5.
10. Add 100 µL per well of **Substrate Solution**. Incubate for 6-7 minutes on microplate shaker at room temperature. **Protect from light.**
11. Add 100 µL per well of **Stop Solution**. The color in the wells should change from blue to yellow. If the color change does not appear uniform, gently tap the plate to ensure thorough mixing.
12. Read plate using a microplate reader set to 450 nm within 15 minutes.

CALCULATION OF RESULTS

Create a standard curve by plotting the log of the known concentrations of the standard dilutions (x-axis) versus the log of its corresponding O.D. (y-axis) and draw the best fit line through the points. It is recommended to use computer software capable of generating a log-log curve fit to more accurately quantify the standard dilutions.

If samples have been diluted, the concentration read from the standard curve must be multiplied by the dilution factor.

SPECIFICITY

Protein	Cross-reactivity
Human SPARC (HEK293 derived)	100%
Mouse SPARC	0


The recombinant human SPARC derived from *E. Coli* or sf21 cells may not be detected by this ELISA kit.

TYPICAL STANDARD CURVE

This standard curve is for demonstration only. A new standard curve should be made for each set of samples assayed.

STANDARD (NG/ML)	AVERAGE OD450 (CORRECTED)
Blank	0 (0.099)
0.19	0.029
0.39	0.063
0.78	0.117
1.563	0.220
3.125	0.564
6.25	1.088
12.5	1.996

SUMMARY OF ASSAY PROCEDURE



nm within 15 minutes.