

MOUSE TNFRSF19 /TROY ELISA KIT

FOR THE QUANTITATIVE DETERMINATION OF
MOUSE SOLUBLE TNFRSF19
CONCENTRATIONS IN SERUM



ALWAYS REFER TO LOT SPECIFIC PROTOCOL
PROVIDED WITH EACH KIT FOR
INSTRUCTIONS. PROTOCOL MUST BE
READ BEFORE USING THIS PRODUCT.

FOR RESEARCH USE ONLY. NOT FOR USE IN
DIAGNOSTIC PROCEDURES.

PRODUCT INFORMATION:

ELISA NAME	MOUSE TNFRSF19 ELISA
Catalog No.	SK00218-03
Formulation	96 T
Lot No.	
Standard range	15.6-1000 pg/mL
Sensitivity	7 pg/mL
Sample Volume	100 µL
Dilution Factor	Optimal dilutions should be determined by each laboratory for each application
Sample Type	Serum
Specificity	Mouse soluble TNFRSF19
Calibration	Mouse soluble TNFRSF19 rec.
Intra-assay Precision	4 - 6%
Inter-assay Precision	8 - 10%
Storage	2 – 8° C
This kit contains sufficient materials to run 35 samples duplicated provided that assay is run according to protocol.	

Order Contact:
AVISCIERA BIOSCIENCE, INC.
 2348 Walsh Ave., Suite C
 Santa Clara, CA 95051
 USA
 Tel: (408) 982 0300
 Fax: (408) 982 0301
 Email: Sales@AvisceraBioscience.com
Info@AvisceraBioscience.com
www.AvisceraBioscience.com

DESCRIPTION

This Mouse Soluble TNFRSF19/TROY ELISA Kit contains the necessary components required for the quantitative measurement of recombinant and/or natural mouse TNFRSF19 from serum in a sandwich ELISA format.

This immunoassay contains recombinant mouse TNFRSF19 and Antibody raised against this protein. Results from this immunoassay have shown to accurately quantify recombinant and natural mouse TNFRSF19 samples.

ASSAY OVERVIEW

This assay employs the quantitative sandwich ELISA format. The plate is pre-coated with a monoclonal antibody specific for mouse TNFRSF19. The capture antibody can bind to the mouse soluble TNFRSF19 in the standard and samples. After washing the plate of any unbound substances, another monoclonal antibody HRP conjugate against mouse TNFRSF19 is added to the wells. Following a wash to remove any unbound antibody enzyme, a substrate solution is added to the wells and color develops in direct proportion to the amount of mouse TNFRSF19 bound in the standard solutions or samples. A standard curve can be established and sample values can be read off the standard curve.

PROCEDURAL LIMITATIONS

_FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.

_This ELISA kit should not be used beyond the expiration date on the kit label.

_Do not mix reagents with those from other lots or sources.

_It is important that the Dilution Buffer selected for the standard curve be consistent with the samples being assayed.

_Each laboratory must determine the optimal dilution factors for the samples being assayed with a pretest. If samples generate values that are not within the dynamic range of the standard curve, further concentrate or dilute the samples as required with Dilution Buffer and repeat the assay.

_Any modifications in buffers, pipetting technique, washing technique, incubation time or temperature, as well as kit age can cause a change in signal.

_Not all interfering factors have been tested in the immunoassay, therefore the possibility of interference cannot be excluded.

COMPONENTS PROVIDED

DESCRIPTION	CODE	QUANTITY
TNFRSF19 Microplate – 96 well microplate coated with an monoclonal antibody specific for mouse TNFRSF19.	218-03-01	1 plate
TNFRSF19 Standard – 1000 pg/vial of lyophilized recombinant mouse TNFRSF19.	218-03-02	1 vial
Detection Antibody – 1.05 mL/vial of 10-fold concentrated solution of Anti mouse TNFRSF19 monoclonal antibody HRP conjugate	218-03-03	1 vial
Positive Control – one vial of lyophilized recombinant Mouse TNFRSF19.	218-03-04	1 vial
Dilution Buffer – 60 mL of buffered solution with preservative.	DB10	1 bottle
Wash Buffer – 50 mL of 10-fold concentrated buffered surfactant with preservative.	WB01	1 bottle
Substrate Solution – 11 mL of substrate solution.	SS01	1 bottle
Stop Solution – 11 mL of 0.5M HCl.	S-STOP	1 bottle
Plate Sealer	EAPS	1 piece
Plastic Pouch	P01	1 piece

STORAGE

Unopened Kit: Store at 2 – 8° C for up to 6 months.

For longer storage, unopened Standard and Positive Control should be stored at -20° C or -70° C.

Detection Antibody-HRP Conjugate 10-fold concentrated solution should be stored at 2 – 8° C (protect from light). Substrate Solution can be stored at 2 – 8° C for up to 6 months (**DO NOT FREEZE** and **PROTECT FROM LIGHT**). All other components can be stored at 2 – 8° C for up to 6 months. Do not use kit past expiration date.

Opened / Reconstituted Reagents: Reconstituted Standard (stock) solution SHOULD BE STORED at -20° C or -70° C for up to one month.

Microplate Wells: Return unused strips to the plastic pouch with the desiccant pack. Microplate may be stored for up to 6 months at 2 – 8° C.

ADDITIONAL MATERIALS REQUIRED

- Microplate reader capable of absorbance measurement at 450 nm.
- Microplate shaker (250 – 300 rpm).
- Microplate washer or manifold dispenser.
- 100 mL and 500 mL graduated cylinders.
- Multi-channel Pipette, Pipettes and pipette tips.
- Deionized or distilled water.

PRECAUTION

This kit should be handled by those persons who have been trained in and can follow the principles of good laboratory practice. Wear protective clothing such as laboratory overalls, safety glasses and gloves. Care should be taken while handling solutions in this kit to avoid contact with skin or eyes, especially with the stop solution because it contains diluted hydrochloric acid. Wash immediately with water in case of contact on skin or eyes.

SAMPLE COLLECTION AND STORAGE

Serum – Use a serum separator tube (SST). Allow blood to clot for 30 minutes. Centrifuge at 1000 x g for 15 minutes and collect serum. Assay samples immediately or aliquot and store at ≤ -20° C. Avoid repeated freeze-thaw cycles.

Optional: Use Aprotinin (enzyme inhibitor) for ALL sample collection to prevent sample degradation. 0.5 TIU per mL of sample solution.

SAMPLE PREPARATION

Optimal dilutions should be determined by each laboratory for each application with a pretest. Use polypropylene test tubes.

REAGENT PREPARATION


Bring all reagents to room temperature before use.

Wash Buffer – If crystals have formed in the concentrate, warm bottle in a water bath until the crystals have completely dissolved. Dilute 50 mL of Wash Buffer Concentrate into 450 mL distilled or deionized water to make 500 mL of 1x Wash Buffer.

Mouse TNFRSF19 Standard – Reconstitute the mouse TNFRSF19 standard with 1.0 mL of Dilution Buffer. The concentration of the reconstituted stock solution is 1000 pg/mL. Allow the stock standard to sit for at least 15 minutes with gentle agitation until

completely dissolved prior to making standard dilutions (see below). Mix each tube thoroughly before the next transfer.

TUBE	STANDARD	DILUTION BUFFER	CONCENTRATION
Stock	powder	1.0 mL	1000 pg/mL
# 1	250µL of stock	250µL	500 pg/mL
# 2	250µL of 1	250µL	250 pg/mL
# 3	250µL of 2	250µL	125 pg/mL
# 4	250µL of 3	250µL	62.5 pg/mL
# 5	250µL of 4	250µL	31.25 pg/mL
# 6	250µL of 5	250µL	15.6 pg/mL



Positive Control - Reconstitute the Positive Control with 1 mL Dilution Buffer. **Note:** Positive Control could be used within a few days if stored at -20° C or -70° C.

Detection Antibody- Pipette 9.45 mL of Dilution Buffer into a 15 mL centrifuge tube and transfer 1.05 mL of 10-fold concentrated Detection Antibody – HRP stock solution to prepare working solution. **Note:** 1x working solution of Detection Antibody - HRP should be used within a few days (**protect from light**). **DO NOT FREEZE.**

ELISA PROTOCOL

Bring all reagents and samples to room temperature before the start of the assay. Blank, standard dilutions, positive control and samples should be assayed in duplicate. ELISA Protocol may need further optimization.

1. Prepare all reagents, standard dilutions, positive control and samples as directed previously.
2. Remove unneeded microplate strips from the plate frame and return them to the plastic pouch with the desiccant pack.
3. Add 100 µL per well of **Dilution Buffer** to Blank wells (A2,A3).
4. Add 100 µL per well of **Standard Dilutions** in reverse order of serial dilution from #6-S (B2, B3 to H2, H3), **sample**, or **positive control** (G4, G5). Cover with plate sealer and incubate for 2 hours on microplate shaker at room temperature.
5. Aspirate and wash each well with 300 µL of **1x Wash Buffer** four times. After the last wash, aspirate any remaining 1x Wash Buffer, invert the plate and blot against clean paper towel(s).
6. Add 100 µL per well of **Detection Antibody working solution**. Cover with plate sealer and incubate for 1 hour on microplate shaker at room temperature. **Protect from light.**
7. Repeat the aspiration and wash as in step 5.
8. Add 100 µL per well of **Substrate Solution**. Incubate for 10-15 minutes on microplate shaker at room temperature. **Protect from light.**
11. Add 100 µL per well of **Stop Solution**. The color in the wells should change from blue to yellow. If the color change does not appear uniform, gently tap the plate to ensure thorough mixing.
12. Read plate using a microplate reader set to 450 nm within 15 minutes.

CALCULATION OF RESULTS

Create a standard curve by plotting the log of the known concentrations of the standard dilutions (x-axis) versus the log of its corresponding O.D. (y-axis) and draw the best fit line through the points. It is recommended to use computer software capable of generating a log-log curve fit to more accurately quantify the standard dilutions.

If samples have been diluted, the concentration read from the standard curve must be multiplied by the dilution factor.

SPECIFICITY

PROTEIN	CROSS-REACTIVITY
Mouse TNFRSF19	100%
Mouse TNFR1	0

TYPICAL STANDARD CURVE

This standard curve is for demonstration only. A new standard curve should be made for each set of samples assayed.

STANDARD (PG/ML)	AVERAGE OD450 (CORRECTED)
Blank	0 (0.108)
15.6	0.041
31.25	0.082
62.5	0.159
125	0.325
250	0.667
500	1.291
1000	2.236

SUMMARY OF ASSAY PROCEDURE

