

# MOUSE SOLUBLE RECEPTOR FOR ADVANCED GLYCOSYLATION END PRODUCTS (sRAGE) ELISA KIT

FOR THE QUANTITATIVE DETERMINATION OF  
MOUSE SOLUBLE RAGE CONCENTRATIONS IN  
PLASMA AND SERUM



ALWAYS REFER TO LOT SPECIFIC PROTOCOL  
PROVIDED WITH EACH KIT FOR  
INSTRUCTIONS. PROTOCOL MUST BE  
READ BEFORE USING THIS PRODUCT.

FOR RESEARCH USE ONLY. NOT FOR USE IN  
DIAGNOSTIC PROCEDURES.

## PRODUCT INFORMATION:

**THIS KIT IS FOR ONE TIME USE ONLY.**

ELISA NAME	SOLUBLE RAGE (MOUSE) ELISA KIT
Catalog No.	SK00112-06
Formulation	96 T
Lot No.	
Standard range	9.7 – 2500 pg/mL
Sensitivity	5 pg/mL
Sample Volume	100 µL
Dilution Factor	2-8 for plasma (Optimal dilutions should be determined by each laboratory for each application)
Sample Type	Serum, Plasma
Specificity	Mouse
Calibration	Recombinant mouse sRAGE (HEK293 derived)
Intra-assay Precision	4 - 6%
Inter-assay Precision	8 - 10%
Storage	2 – 8 °C for 1 month. See page 3 for detail
This kit contains sufficient materials to run 35 samples duplicated provided that assay is run according to protocol.	

Order Contact:  
**AVISCIERA BIOSCIENCE, INC.**  
 2348 Walsh Ave., Suite C  
 Santa Clara, CA 95051  
 Tel: (408) 982 0300  
 Fax: (408) 982 0301  
 Email: [Sales@AvisceraBioscience.com](mailto:Sales@AvisceraBioscience.com)  
[Info@AvisceraBioscience.com](mailto:Info@AvisceraBioscience.com)  
 Website: [www.AvisceraBioscience.com](http://www.AvisceraBioscience.com)  
[www.AvisceraBioscience.net](http://www.AvisceraBioscience.net)

**DESCRIPTION**

This Soluble RAGE (Mouse) ELISA Kit contains the necessary components required for the quantitative measurement of mouse recombinant soluble RAGE (HEK293 cells derived) and/or natural the sRAGE from serum and plasma in a sandwich ELISA format.

This immunoassay contains recombinant glycosylated mouse soluble RAGE (HEK293 cells derived) and antibodies raised, selected and validated by this protein. Results from this immunoassay have shown to accurately quantify recombinant and natural the mouse sRAGE in samples.

**ASSAY OVERVIEW**

This assay employs the quantitative sandwich ELISA format. The plate is pre-coated with an antibody specific for mouse sRAGE. The capture antibody can bind to the mouse sRAGE in the standard and samples. After washing the plate of any unbound substances, a biotinylated antibody against mouse sRAGE is added to the wells. After another washing of the plate, the streptavidin-HRP Conjugate is added. After the last wash to remove any unbound enzyme, a substrate solution (TMB) is added to the wells and color develops in direct proportion to the amount of mouse sRAGE bound in the standard solutions or samples. A standard curve can be established and sample values can be read off the standard curve.

**PROCEDURAL LIMITATIONS**

\_FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.

\_This ELISA kit should not be used beyond the expiration date on the kit label.

\_Do not mix reagents with those from other lots or sources.

\_It is important that the Dilution Buffer selected for the standard curve be consistent with the samples being assayed.

\_Each laboratory must determine the optimal dilution factors for the samples being assayed with a pretest. If samples generate values that are not within the dynamic range of the standard curve, further concentrate or dilute the samples as required with Dilution Buffer and repeat the assay.

\_Any modifications in buffers, pipetting technique, washing technique, incubation time or temperature, as well as kit age can cause a change in signal.

\_Not all interfering factors have been tested in the immunoassay, therefore the possibility of interference cannot be excluded.

**COMPONENTS PROVIDED**

DESCRIPTION	CODE	QUANTITY
<b>Mouse sRAGE Microplate</b> – 96 well microplate coated with antibody specific for the mouse sRAGE.	<b>112-06-01</b>	<b>1 plate</b>
<b>Mouse sRAGE Standard</b> –refer to lot of lyophilized recombinant Mouse sRGAE (HEK293 cells).	<b>112-06-02</b>	<b>1 vial</b>
<b>Detection Antibody Concentrate</b> – refer to lot of 10-fold concentrate of lyophilized biotinylated antibody against human Irisin.	<b>112-06-03</b>	<b>1 vial</b>
<b>Positive Control</b> – one vial of lyophilized recombinant Mouse sRGAE (HEK293 cells).	<b>170-06-04</b>	<b>1 vial</b>
<b>Streptavidin-HRP Conjugate</b> – 120 µL/vial of 100-fold concentrated solution of Streptavidin-HRP conjugate.	<b>SAHRP</b>	<b>1 vial</b>
<b>Dilution Buffer</b> – 40 mL of buffered solution with preservative.	<b>DB01</b>	<b>1 bottle</b>
<b>Antibody Diluent Solution</b> – 12 mL of buffered solution with preservative.	<b>DB12</b>	<b>1 bottle</b>
<b>HRP Diluent Solution</b> – 12 mL of buffered solution with preservative.	<b>DB08C</b>	<b>1 bottle</b>
<b>Wash Buffer</b> – 50 mL of 10-fold concentrated buffered surfactant with preservative.	<b>WB01</b>	<b>1 bottle</b>
<b>TMB Substrate Solution</b> – 11 mL of TMB substrate solution.	<b>TMB01</b>	<b>1 bottle</b>
<b>Stop Solution</b> – 11 mL of 0.5M HCl.	<b>S-STOP</b>	<b>1 bottle</b>
<b>Plate Sealer</b>	<b>EAPS</b>	<b>1 piece</b>
<b>Plastic Pouch</b>	<b>P01</b>	<b>1 piece</b>

**STORAGE**

**Unopened Kit:** Store at 2 – 8 °C for up to 1 month. For longer storage for up to 10 months, unopened Standard, Positive Control, Detection Antibody Concentrate, Dilution Buffer, Antibody Diluent Solution and HRP Diluent Solution should be stored at -20 °C. Streptavidin-HRP Conjugate concentrate and TMB Substrate Solution can be stored only at 2 – 8 °C. Do not use kit past expiration date.

**ADDITIONAL MATERIALS REQUIRED**

- Microplate reader capable of absorbance measurement at 450 nm.
- Microplate shaker (350 – 400 rpm).
- Microplate washer or manifold dispenser.
- 100 mL and 500 mL graduated cylinders.
- Multi-channel Pipette, Pipettes and pipette tips.
- Deionized or distilled water.

**PRECAUTION**

This kit should be handled by those persons who have been trained in and can follow the principles of good laboratory practice. Wear protective clothing such as laboratory overalls, safety glasses and gloves. Care should be taken while handling solutions in this kit to avoid contact with skin or eyes, especially with the stop solution because it contains diluted hydrochloric acid. Wash immediately with water in case of contact on skin or eyes.

**SAMPLE COLLECTION AND STORAGE**

**Plasma** - Collect plasma using EDTA, heparin, or citrate as an anticoagulant. Centrifuge for 15 minutes at 1000x g within 30 minutes of collection. Assay immediately or aliquot and store samples at ≤ -20°C. Avoid repeated freeze-thaw cycles.

**Serum** - Use a serum separator tube (SST) and allow samples to clot for 30 minutes before centrifugation for 15 minutes at 1000 x g. Remove serum and assay immediately or aliquot and store samples at ≤ -20° C. Avoid repeated freeze-thaw cycles.

**Cell Culture Samples - Please use animal free culture media.**

**SAMPLE PREPARATION**

Serum or plasma samples may require at least 2 ~ 8 - fold dilution.

**Optimal dilutions should be determined by each laboratory for each application with a pretest.**

Use polypropylene test tubes.

**REAGENT PREPARATION**

**Bring all reagents to room temperature before use.**

**Wash Buffer** – Dilute 50 mL of Wash Buffer Concentrate into 450 mL distilled or deionized water to make 500 mL of 1x Wash Buffer. If crystals have formed in the concentrate, warm bottle in a water bath until the crystals have completely dissolved.

**Mouse sRAGE Standard** – Reconstitute the human Mouse sRAGE standard with refer to lot of **Dilution Buffer** to this stock standard vial. Allow the stock standard to sit for at least 15 minutes with gentle agitation until completely dissolved prior to making standard dilutions (see below). Mix each tube thoroughly before the next transfer. The **2500 pg/mL** standard serves as the high standard. The Dilution Buffer serves as the zero standard (0 pg/mL).

TUBE	STANDARD	DILUTION BUFFER	CONCENTRATION
Stock	powder	Refer to lot	XXXXX
# 1	Refer to lot	Refer to lot	2500 pg/mL
# 2	150µL of 1	450µL	625 pg/mL
# 3	150µL of 2	450µL	156 pg/mL
# 4	150µL of 3	450µL	39 pg/mL
# 5	150µL of 4	450µL	9.7 pg/mL

**Positive Control** - Reconstitute the Positive Control with refer to lot of **Dilution Buffer** to prepare working solution.

**Detection Antibody** - Reconstitute the Detection Antibody Concentrate with refer to lot of **Antibody Diluent Solution (DB12)** to produce a 10-fold concentrated stock solution. Allow the concentrated solution to sit for at least 5 minutes until completely dissolved. Pipette 9.45 mL of **Antibody Diluent Solution (DB12)** into a 15 mL centrifuge tube and transfer 1.05 mL of 10-fold concentrated stock solution to prepare working solution.

**Streptavidin-HRP Conjugate** - Pipette 10.89 mL of **HRP Diluent Solution (DB08C)** into a 15 mL centrifuge tube and transfer 110 µL of 100-fold concentrated stock solution to prepare working solution (**protect from light**). **DO NOT FREEZE.**

**ELISA PROTOCOL**

**Bring all reagents and samples to room temperature before the start of the assay. Blank, standard dilutions, positive control and samples should be assayed in duplicate. ELISA Protocol may need further optimization.**

1. Prepare all reagents, standard dilutions, positive control and samples as directed previously.
2. Add 100 µL per well of **Dilution Buffer** to Blank wells.
3. Add 100 µL per well of **Standard Dilutions** in reverse order of serial dilution, **samples**, or **positive control**. Cover with plate sealer and incubate for 2 hours on microplate shaker (400-450 rpm) at room temperature.
4. Aspirate and wash each well with 300 µL of **1x Wash Buffer** four times. After the last wash, aspirate any remaining 1x Wash Buffer, invert the plate and blot against clean paper towel(s).
5. Add 100 µL per well of **Detection Antibody working solution**. Cover with plate sealer and incubate for 2 hours on microplate shaker (400-450rpm) at room temperature.
6. Repeat the aspiration and wash as in step 4.
7. Add 100 µL per well of **Streptavidin-HRP Conjugate working solution**. Cover with plate sealer and incubate for 45 minutes on microplate shaker at room temperature. **Protect from light.**
8. Repeat the aspiration and wash as in step 4.
9. Add 100 µL per well of **Substrate Solution**. Incubate for refer to lot on microplate shaker at room temperature. **Protect from light.**
10. Add 100 µL per well of **Stop Solution**. The color in the wells should change from blue to yellow. If the color change does not appear uniform, gently tap the plate to ensure thorough mixing.
11. Read plate using a microplate reader set to 450 nm within 3 minutes.

**CALCULATION OF RESULTS**

Create a standard curve by plotting the log of the known concentrations of the standard dilutions (x-axis) versus the log of its corresponding O.D. (y-axis) and draw the best fit line through the points. It is recommended to use computer software capable of

generating a log-log or 4-parameter curve fit to more accurately quantify the standard dilutions.

If samples have been diluted, the concentration read from the standard curve must be multiplied by the dilution factor.

**SPECIFICITY**

PROTEIN	CROSS-REACTIVITY
sRAGE (HEK293 cells derived)	100%
Human sRAGE (HEK293)	0









The recombinant Mouse sRAGE his tag derived from E. Coli MAY NOT be detected by this ELISA Kit.

**TYPICAL STANDARD CURVE**

This standard curve is for demonstration only. A new standard curve should be made for each set of samples assayed.

STANDARD (PG/ML)	AVERAGE OD450nm (CORRECTED)
Blank	0 (refer to lot)
9.7	0.045
39	0.218
156	0.727
625	1.508
2500	1.959

**SUMMARY OF ASSAY PROCEDURE**

<b>PREPARE REAGENTS, SAMPLES AND STANDARD DILUTIONS</b>

Add 100 µL of standard dilutions, samples and positive control. Cover with plate sealer and incubate 2 hours on microplate shaker at RT.

Aspirate and wash 4 times.

Add 100 µL per well of Detection Antibody working solution. Cover with plate sealer and incubate 2 hours on microplate shaker at RT.

Aspirate and wash 4 times.

Add 100 µL per well of Streptavidin-HRP Conjugate working solution. Cover with plate sealer and incubate 60 minutes on microplate shaker at RT. <b>Protect from light.</b>

Aspirate and wash 4 times.

Add 100 µL per well of Substrate Solution. Incubate refer to lot on microplate shaker at RT. <b>Protect from light.</b>

Add 100 µL per well of Stop Solution. Read at 450 nm within 3 minutes.